

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091922\
 Data File : VD074358.D
 Acq On : 19 Sep 2022 20:52
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 09/20/2022
 Supervised By :Mahesh Dadoda 09/20/2022

Quant Time: Sep 19 23:03:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	89963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	146432	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	133824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	63558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	41973	47.900	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	95.800%		
35) Dibromofluoromethane	7.804	113	45660	47.345	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	94.680%		
50) Toluene-d8	10.269	98	182008	54.695	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	109.400%		
62) 4-Bromofluorobenzene	12.569	95	55366	48.977	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	97.960%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.928	85	34962	42.187	ug/l	93
3) Chloromethane	2.140	50	48900	62.047	ug/l	96
4) Vinyl Chloride	2.275	62	35310	45.651	ug/l	92
5) Bromomethane	2.675	94	12140	25.388	ug/l	99
6) Chloroethane	2.828	64	21817	48.009	ug/l	92
7) Trichlorofluoromethane	3.164	101	66944	44.415	ug/l	98
8) Diethyl Ether	3.587	74	22263	46.337	ug/l	99
9) 1,1,2-Trichlorotrifluo...	3.963	101	45630	48.147	ug/l	99
10) Methyl Iodide	4.158	142	41742	35.656	ug/l	98
11) Tert butyl alcohol	5.052	59	19263	325.294	ug/l #	83
12) 1,1-Dichloroethene	3.934	96	45519	47.184	ug/l	98
13) Acrolein	3.793	56	8665	135.682	ug/l	96
14) Allyl chloride	4.552	41	64272	46.034	ug/l	99
15) Acrylonitrile	5.252	53	48688	229.906	ug/l	98
16) Acetone	4.022	43	34823	176.397	ug/l #	87
17) Carbon Disulfide	4.258	76	124037	39.486	ug/l	100
18) Methyl Acetate	4.563	43	23685	43.684	ug/l	100
19) Methyl tert-butyl Ether	5.316	73	103196	49.952	ug/l	100
20) Methylene Chloride	4.793	84	62734	51.031	ug/l	93
21) trans-1,2-Dichloroethene	5.299	96	50698	47.463	ug/l	98
22) Diisopropyl ether	6.216	45	150947	51.331	ug/l	98
23) Vinyl Acetate	6.152	43	375877	251.612	ug/l	99
24) 1,1-Dichloroethane	6.110	63	94385	49.865	ug/l	96
25) 2-Butanone	7.081	43	60673	206.346	ug/l	99
26) 2,2-Dichloropropane	7.069	77	75007	48.004	ug/l	97
27) cis-1,2-Dichloroethene	7.075	96	58830	49.711	ug/l	97
28) Bromochloromethane	7.422	49	34468	47.160	ug/l	100
29) Tetrahydrofuran	7.446	42	39094	232.001	ug/l	99
30) Chloroform	7.599	83	95264	51.184	ug/l	85
31) Cyclohexane	7.869	56	79420	43.216	ug/l	94
32) 1,1,1-Trichloroethane	7.793	97	81765	49.800	ug/l	99
36) 1,1-Dichloropropene	8.004	75	72602	49.317	ug/l	99
37) Ethyl Acetate	7.163	43	28540	48.391	ug/l	98
38) Carbon Tetrachloride	7.987	117	59613	45.694	ug/l	91
39) Methylcyclohexane	9.269	83	86624	48.646	ug/l	97
40) Benzene	8.246	78	210390	50.969	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091922\
 Data File : VD074358.D
 Acq On : 19 Sep 2022 20:52
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 09/20/2022
 Supervised By :Mahesh Dadoda 09/20/2022

Quant Time: Sep 19 23:03:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.399	41	16547	51.689	ug/l	93
42) 1,2-Dichloroethane	8.328	62	52793	50.078	ug/l	99
43) Isopropyl Acetate	8.363	43	56396	51.830	ug/l	99
44) Trichloroethene	9.022	130	57883	53.336	ug/l	97
45) 1,2-Dichloropropane	9.304	63	54840	52.826	ug/l	99
46) Dibromomethane	9.393	93	26627	49.865	ug/l	97
47) Bromodichloromethane	9.581	83	70684	53.742	ug/l	95
48) Methyl methacrylate	9.381	41	26411	50.157	ug/l	97
49) 1,4-Dioxane	9.381	88	6151	1052.585	ug/l	96
51) 4-Methyl-2-Pentanone	10.157	43	145718	254.893	ug/l	97
52) Toluene	10.334	92	133530	52.517	ug/l	98
53) t-1,3-Dichloropropene	10.551	75	65524	55.839	ug/l	98
54) cis-1,3-Dichloropropene	10.016	75	79408	52.305	ug/l	99
55) 1,1,2-Trichloroethane	10.728	97	36729	52.218	ug/l	98
56) Ethyl methacrylate	10.593	69	47563	51.620	ug/l	99
57) 1,3-Dichloropropane	10.875	76	63903	50.268	ug/l	96
58) 2-Chloroethyl Vinyl ether	9.869	63	22882	105.835	ug/l	99
59) 2-Hexanone	10.916	43	96283	240.784	ug/l	99
60) Dibromochloromethane	11.075	129	43618	51.784	ug/l	100
61) 1,2-Dibromoethane	11.175	107	34265	50.725	ug/l	98
64) Tetrachloroethene	10.804	164	50877	56.087	ug/l	91
65) Chlorobenzene	11.604	112	141076	51.171	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.675	131	48592	51.325	ug/l	99
67) Ethyl Benzene	11.681	91	262719	51.942	ug/l	99
68) m/p-Xylenes	11.787	106	200279	103.373	ug/l	99
69) o-Xylene	12.116	106	96730	54.632	ug/l	100
70) Styrene	12.128	104	161191	53.454	ug/l	98
71) Bromoform	12.292	173	24530	48.284	ug/l #	98
73) Isopropylbenzene	12.416	105	255119	52.190	ug/l	100
74) N-amyl acetate	12.228	43	52458	48.558	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.669	83	42150	49.627	ug/l	100
76) 1,2,3-Trichloropropane	12.716	75	32661m	55.541	ug/l	
77) Bromobenzene	12.692	156	55555	52.009	ug/l	96
78) n-propylbenzene	12.757	91	312153	51.478	ug/l	100
79) 2-Chlorotoluene	12.839	91	166477	52.058	ug/l	100
80) 1,3,5-Trimethylbenzene	12.898	105	205817	51.492	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.463	75	12279	52.332	ug/l #	100
82) 4-Chlorotoluene	12.939	91	169852	51.217	ug/l	100
83) tert-Butylbenzene	13.157	119	186934	52.739	ug/l	98
84) 1,2,4-Trimethylbenzene	13.204	105	206645	52.549	ug/l	99
85) sec-Butylbenzene	13.339	105	276414	52.552	ug/l	100
86) p-Isopropyltoluene	13.451	119	229171	52.134	ug/l	99
87) 1,3-Dichlorobenzene	13.451	146	114206	52.013	ug/l	98
88) 1,4-Dichlorobenzene	13.528	146	109280	50.924	ug/l	99
89) n-Butylbenzene	13.781	91	221083	51.350	ug/l	99
90) Hexachloroethane	14.045	117	33172	39.983	ug/l	98
91) 1,2-Dichlorobenzene	13.822	146	94700	51.152	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.439	75	5666	44.878	ug/l	98
93) 1,2,4-Trichlorobenzene	15.086	180	62173	49.783	ug/l	100
94) Hexachlorobutadiene	15.192	225	37334	52.618	ug/l	97
95) Naphthalene	15.322	128	111031	47.963	ug/l	99
96) 1,2,3-Trichlorobenzene	15.516	180	53234	49.596	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091922\
Data File : VD074358.D
Acq On : 19 Sep 2022 20:52
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022

Quant Time: Sep 19 23:03:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 17 04:16:54 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.875 min Scan# 1069

Delta R.T. 0.000 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion:168 Resp: 8996

Ion Ratio Lower Upper

168 100





99 53.1 43.0 64.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#2

Dichlorodifluoromethane

Concen: 42.187 ug/l

RT: 1.928 min Scan# 58

Delta R.T. -0.006 min

Lab File: VD074358.D




Acq: 19 Sep 2022 20:52


Tgt Ion: 85 Resp: 34962

Ion Ratio Lower Upper

85 100


87 31.5 17.9 53.7





#3
Chloromethane
Concen: 62.047 ug/l
RT: 2.140 min Scan# 94
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 50 Resp: 48900
Ion Ratio Lower Upper
50 100
52 34.5 25.8 38.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#4
Vinyl Chloride
Concen: 45.651 ug/l
RT: 2.275 min Scan# 117
Delta R.T. -0.006 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52



Tgt Ion: 62 Resp: 35310
Ion Ratio Lower Upper
62 100
64 29.1 27.0 40.4









#5
Bromomethane
Concen: 25.388 ug/l
RT: 2.675 min Scan# 185
Delta R.T. -0.012 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 94 Resp: 12140
Ion Ratio Lower Upper
94 100
96 92.4 74.6 112.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022



#6
Chloroethane
Concen: 48.009 ug/l
RT: 2.828 min Scan# 211
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52


Tgt Ion: 64 Resp: 21817
Ion Ratio Lower Upper
64 100
66 35.3 24.6 36.8





#7
 Trichlorofluoromethane
 Concen: 44.415 ug/l
 RT: 3.164 min Scan# 20
 Delta R.T. -0.005 min
 Lab File: VD074358.D
 Acq: 19 Sep 2022 20:52




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion:101 Resp: 66944
 Ion Ratio Lower Upper
 101 100
 103 62.0 50.6 76.0




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 09/20/2022
 Supervised By :Mahesh Dadoda 09/20/2022



#8
 Diethyl Ether
 Concen: 46.337 ug/l
 RT: 3.587 min Scan# 340
 Delta R.T. -0.006 min
 Lab File: VD074358.D
 Acq: 19 Sep 2022 20:52


Tgt Ion: 74 Resp: 22263
 Ion Ratio Lower Upper
 74 100
 45 95.1 47.9 143.6





#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.147 ug/l
RT: 3.963 min Scan# 404
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:101 Resp: 45630
Ion Ratio Lower Upper
101 100
85 45.4 35.0 52.4
151 74.8 59.7 89.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#10
Methyl Iodide
Concen: 35.656 ug/l
RT: 4.158 min Scan# 437
Delta R.T. -0.005 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52



Tgt Ion:142 Resp: 41742
Ion Ratio Lower Upper
142 100
127 44.7 34.6 51.8
141 14.6 11.7 17.5









#11
Tert butyl alcohol
Concen: 325.294 ug/l
RT: 5.052 min Scan# 591
Delta R.T. -0.011 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 59 Resp: 1926
Ion Ratio Lower Upper
59 100
57 3.4 7.7 11.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022



#12
1,1-Dichloroethene
Concen: 47.184 ug/l
RT: 3.934 min Scan# 399
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52


Tgt Ion: 96 Resp: 45519
Ion Ratio Lower Upper
96 100
61 148.6 122.3 183.5
98 63.1 50.7 76.1





#13
Acrolein
Concen: 135.682 ug/l
RT: 3.793 min Scan# 31
Delta R.T. -0.006 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 56 Resp: 866
Ion Ratio Lower Upper
56 100
55 68.7 57.4 86.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#14
Allyl chloride
Concen: 46.034 ug/l
RT: 4.552 min Scan# 504
Delta R.T. -0.011 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52



Tgt Ion: 41 Resp: 64272
Ion Ratio Lower Upper
41 100
39 66.1 52.8 79.2
76 43.0 32.9 49.3





#15

Acrylonitrile

Concen: 229.906 ug/l

RT: 5.252 min Scan# 61

Delta R.T. -0.005 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 53	Resp: 4868
Ion Ratio	Lower Upper
53 100	
52 84.0	65.3 97.9
51 37.1	29.0 43.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#16

Acetone

Concen: 176.397 ug/l




RT: 4.022 min Scan# 414


Delta R.T. 0.000 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52


Tgt Ion: 43	Resp: 34823
Ion Ratio	Lower Upper
43 100	
58 25.4	26.4 39.6





#17
Carbon Disulfide
Concen: 39.486 ug/l
RT: 4.258 min Scan# 41
Delta R.T. -0.011 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 12403
Ion Ratio Lower Upper
76 100
78 8.9 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022



#18
Methyl Acetate
Concen: 43.684 ug/l
RT: 4.563 min Scan# 506
Delta R.T. 0.006 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52

Tgt Ion: 43 Resp: 23685
Ion Ratio Lower Upper
43 100
74 24.1 19.3 28.9









#19
Methyl tert-butyl Ether
Concen: 49.952 ug/l
RT: 5.316 min Scan# 61
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 73 Resp: 103190
Ion Ratio Lower Upper
73 100
57 22.4 17.8 26.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022



#20
Methylene Chloride
Concen: 51.031 ug/l
RT: 4.793 min Scan# 545
Delta R.T. -0.006 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52


Tgt Ion: 84 Resp: 62734
Ion Ratio Lower Upper
84 100
49 119.2 88.4 132.6
51 34.2 28.0 42.0
86 63.8 47.1 70.7





#21
trans-1,2-Dichloroethene
Concen: 47.463 ug/l
RT: 5.299 min Scan# 611
Delta R.T. -0.011 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 96 Resp: 50698
Ion Ratio Lower Upper
96 100
61 133.0 105.1 157.7
98 61.7 47.4 71.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#22
Diisopropyl ether
Concen: 51.331 ug/l
RT: 6.216 min Scan# 787
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52



Tgt Ion: 45 Resp: 150947
Ion Ratio Lower Upper
45 100
43 50.1 38.9 58.3
87 27.7 21.0 31.4
59 12.4 9.6 14.4









#23
 Vinyl Acetate
 Concen: 251.612 ug/l
 RT: 6.152 min Scan# 777
 Delta R.T. -0.005 min
 Lab File: VD074358.D
 Acq: 19 Sep 2022 20:52

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion: 43 Resp: 37587
 Ion Ratio Lower Upper
 43 100
 86 11.2 8.6 13.0




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 09/20/2022
 Supervised By :Mahesh Dadoda 09/20/2022



#24
 1,1-Dichloroethane
 Concen: 49.865 ug/l
 RT: 6.110 min Scan# 769
 Delta R.T. -0.006 min
 Lab File: VD074358.D
 Acq: 19 Sep 2022 20:52


Tgt Ion: 63 Resp: 94385
 Ion Ratio Lower Upper
 63 100
 98 7.3 2.8 8.4
 100 5.6 2.2 6.6





#25
2-Butanone
Concen: 206.346 ug/l
RT: 7.081 min Scan# 91
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 6067
Ion Ratio Lower Upper
43 100
72 27.5 21.8 32.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#26
2,2-Dichloropropane
Concen: 48.004 ug/l
RT: 7.069 min Scan# 932
Delta R.T. -0.006 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Tgt Ion: 77 Resp: 75007
Ion Ratio Lower Upper
77 100
97 24.2 11.5 34.4





#27
 cis-1,2-Dichloroethene
 Concen: 49.711 ug/l
 RT: 7.075 min Scan# 911
 Delta R.T. 0.000 min
 Lab File: VD074358.D
 Acq: 19 Sep 2022 20:52




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion: 96 Resp: 58830
 Ion Ratio Lower Upper
 96 100
 61 139.0 0.0 289.4
 98 65.1 0.0 130.2




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 09/20/2022
 Supervised By :Mahesh Dadoda 09/20/2022



#28
 Bromochloromethane
 Concen: 47.160 ug/l
 RT: 7.422 min Scan# 992
 Delta R.T. 0.000 min
 Lab File: VD074358.D
 Acq: 19 Sep 2022 20:52

Tgt Ion: 49 Resp: 34468
 Ion Ratio Lower Upper
 49 100
 129 1.4 0.0 3.6
 130 77.3 62.2 93.2





#29

Tetrahydrofuran

Concen: 232.001 ug/l

RT: 7.446 min Scan# 995

Delta R.T. 0.006 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42	Resp: 39094
Ion Ratio	Lower Upper
42 100	
72 45.1	36.1 54.1
71 43.3	34.1 51.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#30

Chloroform

Concen: 51.184 ug/l




RT: 7.599 min Scan# 1022


Delta R.T. 0.001 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Tgt Ion: 83	Resp: 95264
Ion Ratio	Lower Upper
83 100	
85 58.6	56.7 85.1





#31

Cyclohexane

Concen: 43.216 ug/l

RT: 7.869 min Scan# 1068

Delta R.T. -0.006 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 56 Resp: 79420





Ion	Ratio	Lower	Upper
56	100		
69	33.0	25.7	38.5
84	85.8	73.9	110.9

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#32

1,1,1-Trichloroethane

Concen: 49.800 ug/l

RT: 7.793 min Scan# 1055




Delta R.T. -0.000 min


Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Tgt Ion: 97 Resp: 81765

Ion	Ratio	Lower	Upper
97	100		
99	63.8	51.2	76.8
61	42.9	33.8	50.8





#33

1,2-Dichloroethane-d4

Concen: 47.900 ug/l

RT: 8.228 min Scan# 1129

Delta R.T. -0.006 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 4197

Ion Ratio Lower Upper

65 100





67 53.8 0.0 110.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1222

Delta R.T. -0.006 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52




Tgt Ion:114 Resp: 146432


Ion Ratio Lower Upper

114 100

63 19.0 0.0 37.6

88 15.6 0.0 28.4





#35

Dibromofluoromethane

Concen: 47.345 ug/l

RT: 7.804 min Scan# 1057

Delta R.T. 0.000 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 113 Resp: 45660





Ion	Ratio	Lower	Upper
113	100		
111	103.5	83.0	124.6
192	19.4	16.2	24.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#36

1,1-Dichloropropene

Concen: 49.317 ug/l

RT: 8.004 min Scan# 1091




Delta R.T. 0.000 min


Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Tgt Ion: 75 Resp: 72602

Ion	Ratio	Lower	Upper
75	100		
110	35.3	17.5	52.5
77	31.8	24.8	37.2





#37

Ethyl Acetate

Concen: 48.391 ug/l

RT: 7.163 min Scan# 949

Delta R.T. -0.006 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 28540





Ion	Ratio	Lower	Upper
43	100		
61	15.6	12.1	18.1
70	9.5	8.5	12.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#38

Carbon Tetrachloride

Concen: 45.694 ug/l

RT: 7.987 min Scan# 1088




Delta R.T. -0.000 min


Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Tgt Ion:117 Resp: 59613

Ion	Ratio	Lower	Upper
117	100		
119	88.5	78.8	118.2
121	27.6	23.9	35.9





#39

Methylcyclohexane

Concen: 48.646 ug/l

RT: 9.269 min Scan# 11306

Delta R.T. -0.006 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83	Resp: 86624
Ion Ratio	Lower Upper
83 100	
55 70.0	57.2 85.8
98 45.3	38.3 57.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#40

Benzene

Concen: 50.969 ug/l




RT: 8.246 min Scan# 1132


Delta R.T. -0.005 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Tgt Ion: 78	Resp: 210390
Ion Ratio	Lower Upper
78 100	
77 24.6	19.4 29.2






#41
Methacrylonitrile
Concen: 51.689 ug/l
RT: 7.399 min Scan# 988
Delta R.T. -0.005 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 41 Resp: 1654
Ion Ratio Lower Upper
41 100
39 52.9 39.8 59.6
67 77.1 56.1 84.1
52 30.4 22.0 33.0

Manual Integrations

APPROVED




Reviewed By :Krupa Patel 09/20/2022


Supervised By :Mahesh Dadoda 09/20/2022



#42
1,2-Dichloroethane
Concen: 50.078 ug/l
RT: 8.328 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52


Tgt Ion: 62 Resp: 52793
Ion Ratio Lower Upper
62 100
98 10.2 0.0 21.4





#43
Isopropyl Acetate
Concen: 51.830 ug/l
RT: 8.363 min Scan# 1152
Delta R.T. 0.006 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 56390
Ion Ratio Lower Upper
43 100
61 31.6 24.5 36.7
87 13.4 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#44
Trichloroethene
Concen: 53.336 ug/l
RT: 9.022 min Scan# 1264
Delta R.T. -0.006 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52



Tgt Ion:130 Resp: 57883
Ion Ratio Lower Upper
130 100
95 98.1 0.0 202.2





#45

1,2-Dichloropropane

Concen: 52.826 ug/l

RT: 9.304 min Scan# 1312

Delta R.T. 0.000 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 54840

Ion Ratio Lower Upper

63 100





65 30.1 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#46

Dibromomethane

Concen: 49.865 ug/l

RT: 9.393 min Scan# 1327

Delta R.T. 0.001 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52




Tgt Ion: 93 Resp: 26627


Ion Ratio Lower Upper

93 100

95 84.0 67.7 101.5


174 94.4 79.6 119.4





#47
Bromodichloromethane
Concen: 53.742 ug/l
RT: 9.581 min Scan# 1359
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 70684
Ion Ratio Lower Upper
83 100
85 62.5 53.4 80.2
127 8.1 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022



#48
Methyl methacrylate
Concen: 50.157 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52

Tgt Ion: 41 Resp: 26411
Ion Ratio Lower Upper
41 100
69 89.0 72.7 109.1
39 49.4 41.6 62.4





#49

1,4-Dioxane

Concen: 1052.585 ug/l

RT: 9.381 min Scan# 11

Delta R.T. -0.000 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 88 Resp: 615

Ion Ratio Lower Upper

88 100

43 30.9 27.0 40.4





58 64.4 54.2 81.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#50

Toluene-d8

Concen: 54.695 ug/l

RT: 10.269 min Scan# 1476

Delta R.T. 0.000 min

Lab File: VD074358.D




Acq: 19 Sep 2022 20:52


Tgt Ion: 98 Resp: 182008

Ion Ratio Lower Upper

98 100


100 65.7 51.8 77.8





#51
4-Methyl-2-Pentanone
Concen: 254.893 ug/l
RT: 10.157 min Scan# 1457
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 145718
Ion Ratio Lower Upper
43 100
58 40.0 33.3 49.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#52
Toluene
Concen: 52.517 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Tgt Ion: 92 Resp: 133530
Ion Ratio Lower Upper
92 100
91 177.5 139.3 208.9





#53
 t-1,3-Dichloropropene
 Concen: 55.839 ug/l
 RT: 10.551 min Scan# 1524
 Delta R.T. 0.000 min
 Lab File: VD074358.D
 Acq: 19 Sep 2022 20:52




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion: 75 Resp: 65524
 Ion Ratio Lower Upper
 75 100
 77 33.2 25.7 38.5

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 09/20/2022
 Supervised By :Mahesh Dadoda 09/20/2022




#54
 cis-1,3-Dichloropropene
 Concen: 52.305 ug/l
 RT: 10.016 min Scan# 1433
 Delta R.T. 0.000 min
 Lab File: VD074358.D
 Acq: 19 Sep 2022 20:52




Tgt Ion: 75 Resp: 79408
 Ion Ratio Lower Upper
 75 100
 77 31.5 25.4 38.0
 39 38.2 31.6 47.4





#55
1,1,2-Trichloroethane
Concen: 52.218 ug/l
RT: 10.728 min Scan# 1554
Delta R.T. -0.006 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:	97	Resp:	36729
Ion Ratio	100	Lower	Upper
97	100		
83	86.5	68.0	102.0
85	56.6	45.8	68.8
99	66.4	50.7	76.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#56
Ethyl methacrylate
Concen: 51.620 ug/l
RT: 10.593 min Scan# 1531
Delta R.T. -0.005 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52



Tgt Ion:	69	Resp:	47563
Ion Ratio	100	Lower	Upper
69	100		
41	61.7	49.4	74.2
39	31.9	24.2	36.4





#57

1,3-Dichloropropane

Concen: 50.268 ug/l

RT: 10.875 min Scan# 1111

Delta R.T. -0.000 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 6390

Ion Ratio Lower Upper

76 100





78 33.8 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#58

2-Chloroethyl Vinyl ether

Concen: 105.835 ug/l

RT: 9.869 min Scan# 1408

Delta R.T. 0.000 min

Lab File: VD074358.D




Acq: 19 Sep 2022 20:52


Tgt Ion: 63 Resp: 22882

Ion Ratio Lower Upper

63 100


106 27.0 21.9 32.9





#59
2-Hexanone
Concen: 240.784 ug/l
RT: 10.916 min Scan# 1586
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 96283
Ion Ratio Lower Upper
43 100
58 57.7 28.6 85.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022



#60
Dibromochloromethane
Concen: 51.784 ug/l
RT: 11.075 min Scan# 1613
Delta R.T. 0.006 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52


Tgt Ion:129 Resp: 43618
Ion Ratio Lower Upper
129 100
127 77.5 39.0 116.9





#61
1,2-Dibromoethane
Concen: 50.725 ug/l
RT: 11.175 min Scan# 1630
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:107 Resp: 3426
Ion Ratio Lower Upper
107 100
109 94.4 73.7 110.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#62
4-Bromofluorobenzene
Concen: 48.977 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Tgt Ion: 95 Resp: 55366
Ion Ratio Lower Upper
95 100
174 88.9 0.0 179.2
176 87.4 0.0 171.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.575 min Scan# 1698
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:117 Resp: 133824
Ion Ratio Lower Upper
117 100
82 54.1 44.2 66.2
119 30.5 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#64
Tetrachloroethene
Concen: 56.087 ug/l
RT: 10.804 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52



Tgt Ion:164 Resp: 50877
Ion Ratio Lower Upper
164 100
166 125.8 93.8 140.8
129 96.3 68.3 102.5
131 90.1 67.1 100.7





#65

Chlorobenzene

Concen: 51.171 ug/l

RT: 11.604 min Scan# 1703

Delta R.T. 0.000 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion:112 Resp: 141070

Ion Ratio Lower Upper

112 100





114 32.3 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#66

1,1,1,2-Tetrachloroethane

Concen: 51.325 ug/l

RT: 11.675 min Scan# 1715

Delta R.T. -0.000 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52




Tgt Ion:131 Resp: 48592


Ion Ratio Lower Upper

131 100

133 94.9 47.8 143.4

119 66.5 33.1 99.5





#67

Ethyl Benzene

Concen: 51.942 ug/l

RT: 11.681 min Scan# 1716

Delta R.T. -0.000 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 91 Resp: 262719

Ion Ratio Lower Upper

91 100





106 31.4 24.8 37.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#68

m/p-Xylenes

Concen: 103.373 ug/l

RT: 11.787 min Scan# 1734

Delta R.T. 0.001 min

Lab File: VD074358.D




Acq: 19 Sep 2022 20:52


Tgt Ion:106 Resp: 200279

Ion Ratio Lower Upper

106 100


91 199.8 158.6 237.8





#69
o-Xylene
Concen: 54.632 ug/l
RT: 12.116 min Scan# 1790
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion:106 Resp: 96730
Ion Ratio Lower Upper
106 100
91 208.0 104.0 312.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022



#70
Styrene
Concen: 53.454 ug/l
RT: 12.128 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52


Tgt Ion:104 Resp: 161191
Ion Ratio Lower Upper
104 100
78 46.7 38.4 57.6
103 57.0 44.1 66.1





#71
Bromoform
Concen: 48.284 ug/l
RT: 12.292 min Scan# 1820
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:173 Resp: 24530
Ion Ratio Lower Upper
173 100
175 49.6 24.2 72.6
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.510 min Scan# 2027
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52



Tgt Ion:152 Resp: 63558
Ion Ratio Lower Upper
152 100
115 58.6 28.4 85.2
150 176.9 0.0 351.0





#73

Isopropylbenzene

Concen: 52.190 ug/l

RT: 12.416 min Scan# 1841

Delta R.T. -0.000 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC

Tgt Ion:105 Resp: 255119

Ion Ratio Lower Upper

105 100





120 26.3 13.2 39.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#74

N-ethyl acetate

Concen: 48.558 ug/l

RT: 12.228 min Scan# 1809

Delta R.T. -0.000 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Tgt Ion: 43 Resp: 52458




Ion Ratio Lower Upper


43 100

70 44.9 34.6 51.8

55 24.6 18.2 27.2


61 25.0 20.6 31.0





#75
1,1,2,2-Tetrachloroethane
Concen: 49.627 ug/l
RT: 12.669 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 42150
Ion Ratio Lower Upper
83 100
131 9.9 4.9 14.7
85 64.3 32.3 96.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022



#76
1,2,3-Trichloropropane
Concen: 55.541 ug/l m
RT: 12.716 min Scan# 1892
Delta R.T. -0.006 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52


Tgt Ion: 75 Resp: 32661
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 52.009 ug/l
RT: 12.692 min Scan# 1888
Delta R.T. -0.006 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:156 Resp: 5555
Ion Ratio Lower Upper
156 100
77 175.3 91.1 273.3
158 95.8 49.4 148.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#78
n-propylbenzene
Concen: 51.478 ug/l
RT: 12.757 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52



Tgt Ion: 91 Resp: 312153
Ion Ratio Lower Upper
91 100
120 23.0 11.6 34.7









#79
2-Chlorotoluene
Concen: 52.058 ug/l
RT: 12.839 min Scan# 1914
Delta R.T. -0.006 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 16647
Ion Ratio Lower Upper
91 100
126 34.8 17.3 51.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022



#80
1,3,5-Trimethylbenzene
Concen: 51.492 ug/l
RT: 12.898 min Scan# 1923
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52


Tgt Ion:105 Resp: 205817
Ion Ratio Lower Upper
105 100
120 49.8 24.6 74.0





#81
trans-1,4-Dichloro-2-butene
Concen: 52.332 ug/l
RT: 12.463 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 12279
Ion Ratio Lower Upper
75 100
53 81.2 0.0 0.0
89 42.9 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022



#82
4-Chlorotoluene
Concen: 51.217 ug/l
RT: 12.939 min Scan# 1930
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52


Tgt Ion: 91 Resp: 169852
Ion Ratio Lower Upper
91 100
126 34.6 17.2 51.6





#83
tert-Butylbenzene
Concen: 52.739 ug/l
RT: 13.157 min Scan# 1967
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:119 Resp: 186934
Ion Ratio Lower Upper
119 100
91 61.2 31.3 93.9
134 25.7 13.4 40.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#84
1,2,4-Trimethylbenzene
Concen: 52.549 ug/l
RT: 13.204 min Scan# 1975
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Tgt Ion:105 Resp: 206645
Ion Ratio Lower Upper
105 100
120 45.6 23.1 69.3





#85
 sec-Butylbenzene
 Concen: 52.552 ug/l
 RT: 13.339 min Scan# 1998
 Delta R.T. 0.000 min
 Lab File: VD074358.D
 Acq: 19 Sep 2022 20:52




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion:105 Resp: 276414
 Ion Ratio Lower Upper
 105 100
 134 19.7 10.0 29.8

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 09/20/2022
 Supervised By :Mahesh Dadoda 09/20/2022




#86
 p-Isopropyltoluene
 Concen: 52.134 ug/l
 RT: 13.451 min Scan# 1717
 Delta R.T. 0.000 min
 Lab File: VD074358.D
 Acq: 19 Sep 2022 20:52




Tgt Ion:119 Resp: 229171
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.1 39.3
 91 22.6 11.2 33.5





#87
1,3-Dichlorobenzene
Concen: 52.013 ug/l
RT: 13.451 min Scan# 2017
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 114200
Ion Ratio Lower Upper
146 100
111 37.3 18.1 54.4
148 62.2 31.7 95.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022




#88
1,4-Dichlorobenzene
Concen: 50.924 ug/l
RT: 13.528 min Scan# 2030
Delta R.T. 0.001 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Tgt Ion:146 Resp: 109280
Ion Ratio Lower Upper
146 100
111 37.1 18.4 55.2
148 63.8 32.1 96.5





#89
n-Butylbenzene
Concen: 51.350 ug/l
RT: 13.781 min Scan# 2073
Delta R.T. 0.001 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 221083
Ion Ratio Lower Upper
91 100
92 52.7 26.6 79.7
134 25.7 13.1 39.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022



#90
Hexachloroethane
Concen: 39.983 ug/l
RT: 14.045 min Scan# 2118
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52


Tgt Ion:117 Resp: 33172
Ion Ratio Lower Upper
117 100
201 73.8 38.0 113.9





#91
1,2-Dichlorobenzene
Concen: 51.152 ug/l
RT: 13.822 min Scan# 2080
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC



Tgt Ion:146 Resp: 94700
Ion Ratio Lower Upper
146 100
111 39.7 19.4 58.4
148 65.0 31.5 94.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 44.878 ug/l
RT: 14.439 min Scan# 2185
Delta R.T. 0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52

Tgt Ion: 75 Resp: 5666
Ion Ratio Lower Upper
75 100
155 95.3 49.2 147.6
157 121.9 62.1 186.2





#93

1,2,4-Trichlorobenzene

Concen: 49.783 ug/l

RT: 15.086 min Scan# 2295

Delta R.T. 0.000 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	180	Resp:	6217
Ion Ratio	Lower	Upper	
180	100		
182	95.3	47.6	142.9
145	28.8	14.4	43.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/20/2022

Supervised By :Mahesh Dadoda 09/20/2022



#94

Hexachlorobutadiene

Concen: 52.618 ug/l




RT: 15.192 min Scan# 2313


Delta R.T. 0.000 min

Lab File: VD074358.D

Acq: 19 Sep 2022 20:52


Tgt Ion:	225	Resp:	37334
Ion Ratio	Lower	Upper	
225	100		
223	62.6	32.4	97.0
227	62.5	32.7	98.1





#95
Naphthalene
Concen: 47.963 ug/l
RT: 15.322 min Scan# 2335
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC




Tgt Ion:128 Resp: 11103
Ion Ratio Lower Upper
128 100
127 14.1 10.9 16.3
129 11.1 8.8 13.2

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 09/20/2022
Supervised By :Mahesh Dadoda 09/20/2022



#96
1,2,3-Trichlorobenzene
Concen: 49.596 ug/l
RT: 15.516 min Scan# 2368
Delta R.T. -0.000 min
Lab File: VD074358.D
Acq: 19 Sep 2022 20:52



Tgt Ion:180 Resp: 53234
Ion Ratio Lower Upper
180 100
182 98.7 47.5 142.5
145 31.8 15.8 47.3

