

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092622\
 Data File : VD074420.D
 Acq On : 26 Sep 2022 15:06
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Sep 27 03:23:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D092622S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 03:19:14 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	33644	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	94531	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	102982	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	50691	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	24041	82.999	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 166.000%#		
35) Dibromofluoromethane	7.804	113	35046	55.446	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 110.900%		
50) Toluene-d8	10.269	98	118394	55.073	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 110.140%		
62) 4-Bromofluorobenzene	12.569	95	42984	58.250	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 116.500%		
Target Compounds						
20) Methylene Chloride	4.793	84	10355m	20.037	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092622\
Data File : VD074420.D
Acq On : 26 Sep 2022 15:06
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Sep 27 03:23:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D092622S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 03:19:14 2022
Response via : Initial Calibration

