

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092622\
 Data File : VD074431.D
 Acq On : 26 Sep 2022 20:55
 Operator : VA/SY
 Sample : N4803-01
 Misc : 5.60G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SASP2-T-1-(10-17)

Quant Time: Sep 27 04:11:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D092622S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 03:19:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	62653	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	103838	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	99148	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	49398	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	25754	47.745	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.500%
35) Dibromofluoromethane	7.804	113	35133	50.284	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.560%
50) Toluene-d8	10.269	98	120513	51.034	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.060%
62) 4-Bromofluorobenzene	12.569	95	38570	46.347	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.700%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092622\
Data File : VD074431.D
Acq On : 26 Sep 2022 20:55
Operator : VA/SY
Sample : N4803-01
Misc : 5.60G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-1-(10-17)

Quant Time: Sep 27 04:11:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D092622S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 03:19:14 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1129
Delta R.T. 0.000 min
Lab File: VD074431.D
Acq: 26 Sep 2022 20:55




Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-1-(10-17)


Tgt Ion:168 Resp: 62653
Ion Ratio Lower Upper
168 100
99 49.8 40.9 61.3



#33
1,2-Dichloroethane-d4
Concen: 47.745 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. -0.006 min
Lab File: VD074431.D
Acq: 26 Sep 2022 20:55

Tgt Ion: 65 Resp: 25754
Ion Ratio Lower Upper
65 100
67 55.0 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 11

Delta R.T. 0.000 min

Lab File: VD074431.D

Acq: 26 Sep 2022 20:55

Instrument :

MSVOA_D

ClientSampleId :

SASP2-T-1-(10-17)





Tgt Ion:114 Resp: 103838

Ion Ratio Lower Upper

114 100

63 15.6 0.0 33.2

88 15.1 0.0 28.2



#35

Dibromofluoromethane

Concen: 50.284 ug/l

RT: 7.804 min Scan# 1057

Delta R.T. -0.006 min

Lab File: VD074431.D

Acq: 26 Sep 2022 20:55




Tgt Ion:113 Resp: 35133


Ion Ratio Lower Upper

113 100

111 105.5 82.3 123.5

192 19.4 16.4 24.6









#50
Toluene-d8
Concen: 51.034 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD074431.D
Acq: 26 Sep 2022 20:55




Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-1-(10-17)


Tgt Ion: 98 Resp: 120513
Ion Ratio Lower Upper
98 100
100 62.7 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 46.347 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VD074431.D
Acq: 26 Sep 2022 20:55

Tgt Ion: 95 Resp: 38570
Ion Ratio Lower Upper
95 100
174 98.5 0.0 197.0
176 95.9 0.0 182.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. 0.000 min
Lab File: VD074431.D
Acq: 26 Sep 2022 20:55

Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-1-(10-17)

Tgt Ion:117 Resp: 99148
Ion Ratio Lower Upper
117 100
82 51.7 41.6 62.4
119 30.6 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. 0.006 min
Lab File: VD074431.D
Acq: 26 Sep 2022 20:55

Tgt Ion:152 Resp: 49398
Ion Ratio Lower Upper
152 100
115 54.4 26.8 80.4
150 157.1 0.0 348.2

