

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092622\
 Data File : VD074432.D
 Acq On : 26 Sep 2022 21:24
 Operator : VA/SY
 Sample : N4803-02
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SASP2-T-1-(17-23)

Quant Time: Sep 27 04:12:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D092622S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 03:19:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	63547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	104899	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	99330	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	50074	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	27126	49.581	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.160%
35) Dibromofluoromethane	7.810	113	36438	51.722	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.440%
50) Toluene-d8	10.269	98	122923	51.528	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.060%
62) 4-Bromofluorobenzene	12.569	95	40610	48.590	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.180%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092622\
Data File : VD074432.D
Acq On : 26 Sep 2022 21:24
Operator : VA/SY
Sample : N4803-02
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-1-(17-23)

Quant Time: Sep 27 04:12:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D092622S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 03:19:14 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD074432.D
Acq: 26 Sep 2022 21:24




Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-1-(17-23)


Tgt Ion:168 Resp: 63547
Ion Ratio Lower Upper
168 100
99 48.8 40.9 61.3



#33
1,2-Dichloroethane-d4
Concen: 49.581 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VD074432.D
Acq: 26 Sep 2022 21:24

Tgt Ion: 65 Resp: 27126
Ion Ratio Lower Upper
65 100
67 55.4 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1

Delta R.T. 0.000 min

Lab File: VD074432.D

Acq: 26 Sep 2022 21:24

Instrument :

MSVOA_D

ClientSampleId :

SASP2-T-1-(17-23)





Tgt Ion:114 Resp: 104899

Ion Ratio Lower Upper

114 100

63 16.4 0.0 33.2

88 14.1 0.0 28.2



#35

Dibromofluoromethane

Concen: 51.722 ug/l

RT: 7.810 min Scan# 1058

Delta R.T. 0.000 min

Lab File: VD074432.D

Acq: 26 Sep 2022 21:24




Tgt Ion:113 Resp: 36438


Ion Ratio Lower Upper

113 100

111 102.8 82.3 123.5

192 20.2 16.4 24.6









#50
Toluene-d8
Concen: 51.528 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD074432.D
Acq: 26 Sep 2022 21:24




Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-1-(17-23)


Tgt Ion: 98 Resp: 122923
Ion Ratio Lower Upper
98 100
100 63.0 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 48.590 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VD074432.D
Acq: 26 Sep 2022 21:24

Tgt Ion: 95 Resp: 40610
Ion Ratio Lower Upper
95 100
174 97.1 0.0 197.0
176 94.1 0.0 182.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 117
Delta R.T. 0.000 min
Lab File: VD074432.D
Acq: 26 Sep 2022 21:24

Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-1-(17-23)

Tgt Ion:117 Resp: 99330
Ion Ratio Lower Upper
117 100
82 51.1 41.6 62.4
119 33.3 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.510 min Scan# 2027
Delta R.T. 0.000 min
Lab File: VD074432.D
Acq: 26 Sep 2022 21:24

Tgt Ion:152 Resp: 50074
Ion Ratio Lower Upper
152 100
115 54.2 26.8 80.4
150 156.7 0.0 348.2

