



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092622\
 Data File : VD074413.D
 Acq On : 26 Sep 2022 10:27
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D092622S.M
 Title : SW846 8260
 Last Update : Tue Sep 27 03:19:14 2022



AutoFind: Scans 1866, 1867, 1868; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.3	3146	PASS
75	95	30	60	48.7	10012	PASS
95	95	100	100	100.0	20568	PASS
96	95	5	9	6.5	1344	PASS
173	174	0.00	2	1.2	243	PASS
174	95	50	100	97.2	19989	PASS
175	174	5	9	7.9	1573	PASS
176	174	95	101	98.4	19672	PASS
177	176	5	9	7.3	1437	PASS