

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092722\
 Data File : VD074442.D
 Acq On : 27 Sep 2022 13:48
 Operator : VA/SY
 Sample : N4826-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-03-092622

Quant Time: Sep 28 04:40:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D092622S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 03:19:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	65936	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	110410	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	93734	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	44136	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	25114	44.240	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.480%
35) Dibromofluoromethane	7.804	113	32926	43.890	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.780%
50) Toluene-d8	10.269	98	109199	43.491	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	86.980%
62) 4-Bromofluorobenzene	12.569	95	31193	33.636	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	67.280%
Target Compounds						
20) Methylene Chloride	4.793	84	17240	15.808	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092722\
Data File : VD074442.D
Acq On : 27 Sep 2022 13:48
Operator : VA/SY
Sample : N4826-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-03-092622

Quant Time: Sep 28 04:40:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D092622S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 03:19:14 2022
Response via : Initial Calibration

