

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092722\
 Data File : VD074443.D
 Acq On : 27 Sep 2022 14:16
 Operator : VA/SY
 Sample : N4812-01
 Misc : 6.12G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-3

Quant Time: Sep 28 04:40:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D092622S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 03:19:14 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	7357	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	11706	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	12995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	7329	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	4716	74.456	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	148.920%	
35) Dibromofluoromethane	7.804	113	4596	58.933	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	117.860%	
50) Toluene-d8	10.269	98	12011	45.119	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	90.240%	
62) 4-Bromofluorobenzene	12.569	95	4771	51.511	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	103.020%	
Target Compounds						
20) Methylene Chloride	4.799	84	2148	18.593	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092722\
Data File : VD074443.D
Acq On : 27 Sep 2022 14:16
Operator : VA/SY
Sample : N4812-01
Misc : 6.12G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-3

Quant Time: Sep 28 04:40:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D092622S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 03:19:14 2022
Response via : Initial Calibration

