

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101122\
 Data File : VD074519.D
 Acq On : 11 Oct 2022 15:17
 Operator : VA/SY
 Sample : VSTDICC020
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICC020

Quant Time: Oct 12 03:09:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 12 03:06:42 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 10/12/2022
 Supervised By :Mahesh Dadoda 10/13/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	92708	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	161053	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	146903	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	68631	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	12852	19.261	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	38.520%#		
35) Dibromofluoromethane	7.805	113	19799	22.424	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	44.840%#		
50) Toluene-d8	10.269	98	47400	16.377	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	32.760%#		
62) 4-Bromofluorobenzene	12.569	95	29984	28.269	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	56.540%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.934	85	15557	17.396	ug/l	99
3) Chloromethane	2.146	50	22094	19.940	ug/l	88
4) Vinyl Chloride	2.281	62	21475	18.724	ug/l	95
5) Bromomethane	2.670	94	16740	21.489	ug/l	94
6) Chloroethane	2.828	64	15360	20.419	ug/l	99
7) Trichlorofluoromethane	3.170	101	32422	18.693	ug/l	94
8) Diethyl Ether	3.593	74	10137	19.072	ug/l	95
9) 1,1,2-Trichlorotrifluo...	3.958	101	22818	19.447	ug/l	95
10) Methyl Iodide	4.158	142	23876	18.620	ug/l	99
11) Tert butyl alcohol	5.022	59	16269m	130.694	ug/l	
12) 1,1-Dichloroethene	3.934	96	23220	19.468	ug/l	98
13) Acrolein	3.793	56	4584	117.834	ug/l	98
14) Allyl chloride	4.558	41	25275	19.218	ug/l	98
15) Acrylonitrile	5.252	53	20412	97.362	ug/l	97
16) Acetone	4.022	43	14825	83.264	ug/l	100
17) Carbon Disulfide	4.269	76	70455	19.974	ug/l	97
18) Methyl Acetate	4.564	43	11242	20.949	ug/l	98
19) Methyl tert-butyl Ether	5.311	73	46292	20.312	ug/l	94
20) Methylene Chloride	4.799	84	36860	22.369	ug/l	89
21) trans-1,2-Dichloroethene	5.305	96	25253	19.285	ug/l	97
22) Diisopropyl ether	6.222	45	55784	21.220	ug/l	98
23) Vinyl Acetate	6.158	43	108722	92.989	ug/l	98
24) 1,1-Dichloroethane	6.111	63	40572	20.111	ug/l	98
25) 2-Butanone	7.087	43	22479	94.079	ug/l	98
26) 2,2-Dichloropropane	7.075	77	34994	19.545	ug/l	98
27) cis-1,2-Dichloroethene	7.075	96	28909	20.962	ug/l	98
28) Bromochloromethane	7.428	49	12208	23.014	ug/l	86
29) Tetrahydrofuran	7.446	42	13483	102.735	ug/l	97
30) Chloroform	7.593	83	41330	20.754	ug/l	93
31) Cyclohexane	7.881	56	32183	19.097	ug/l	92
32) 1,1,1-Trichloroethane	7.793	97	35099	19.705	ug/l	97
36) 1,1-Dichloropropene	8.010	75	29657	19.035	ug/l	98
37) Ethyl Acetate	7.169	43	10243	20.400	ug/l	98
38) Carbon Tetrachloride	7.987	117	27307	19.057	ug/l	92
39) Methylcyclohexane	9.275	83	37927	19.685	ug/l	98
40) Benzene	8.252	78	90228	19.634	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101122\
 Data File : VD074519.D
 Acq On : 11 Oct 2022 15:17
 Operator : VA/SY
 Sample : VSTDICC020
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 10/12/2022
 Supervised By :Mahesh Dadoda 10/13/2022

Quant Time: Oct 12 03:09:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 12 03:06:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.399	41	5353	17.712	ug/l #	94
42) 1,2-Dichloroethane	8.322	62	20880	19.294	ug/l	99
43) Isopropyl Acetate	8.363	43	18220	19.596	ug/l #	83
44) Trichloroethene	9.028	130	25842	19.857	ug/l	95
45) 1,2-Dichloropropane	9.305	63	20865	19.941	ug/l	91
46) Dibromomethane	9.393	93	11634	19.967	ug/l	95
47) Bromodichloromethane	9.581	83	29114	20.090	ug/l	92
48) Methyl methacrylate	9.375	41	8995	20.102	ug/l	95
49) 1,4-Dioxane	9.387	88	2650	452.160	ug/l #	92
51) 4-Methyl-2-Pentanone	10.157	43	46875	102.177	ug/l	95
52) Toluene	10.334	92	58397	19.984	ug/l	97
53) t-1,3-Dichloropropene	10.552	75	25522	19.882	ug/l	95
54) cis-1,3-Dichloropropene	10.016	75	31832	19.750	ug/l	97
55) 1,1,2-Trichloroethane	10.734	97	16262	20.709	ug/l	94
56) Ethyl methacrylate	10.593	69	18348	20.520	ug/l	96
57) 1,3-Dichloropropane	10.881	76	27515	21.211	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.869	63	13962	75.624	ug/l	99
59) 2-Hexanone	10.916	43	31831	103.458	ug/l	99
60) Dibromochloromethane	11.075	129	19695	20.776	ug/l	98
61) 1,2-Dibromoethane	11.181	107	14917	20.446	ug/l	100
64) Tetrachloroethene	10.810	164	21499	19.753	ug/l	89
65) Chlorobenzene	11.604	112	63736	20.147	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.681	131	20872	19.572	ug/l	93
67) Ethyl Benzene	11.681	91	109282	19.491	ug/l	98
68) m/p-Xylenes	11.793	106	88638	39.706	ug/l	97
69) o-Xylene	12.116	106	41300	19.954	ug/l	98
70) Styrene	12.134	104	71802	20.476	ug/l	99
71) Bromoform	12.293	173	10775	20.210	ug/l #	97
73) Isopropylbenzene	12.416	105	108999	19.984	ug/l	100
74) N-amyl acetate	12.228	43	17064	19.990	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.669	83	17498	20.217	ug/l	97
76) 1,2,3-Trichloropropane	12.716	75	12409m	24.108	ug/l	
77) Bromobenzene	12.693	156	24542	19.816	ug/l	97
78) n-propylbenzene	12.757	91	133711	20.204	ug/l	99
79) 2-Chlorotoluene	12.845	91	71733	20.297	ug/l	100
80) 1,3,5-Trimethylbenzene	12.898	105	88409	19.919	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.463	75	4183	17.553	ug/l	96
82) 4-Chlorotoluene	12.945	91	72484	20.218	ug/l	97
83) tert-Butylbenzene	13.163	119	76708	19.770	ug/l	98
84) 1,2,4-Trimethylbenzene	13.204	105	87833	19.988	ug/l	98
85) sec-Butylbenzene	13.340	105	116816	19.778	ug/l	99
86) p-Isopropyltoluene	13.457	119	96438	19.559	ug/l	99
87) 1,3-Dichlorobenzene	13.451	146	50058	19.876	ug/l	99
88) 1,4-Dichlorobenzene	13.534	146	49164	19.737	ug/l	98
89) n-Butylbenzene	13.781	91	91657	19.895	ug/l	98
90) Hexachloroethane	14.045	117	16834	19.237	ug/l	92
91) 1,2-Dichlorobenzene	13.828	146	42607	19.778	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.440	75	2406	19.758	ug/l	85
93) 1,2,4-Trichlorobenzene	15.092	180	27107	19.121	ug/l	97
94) Hexachlorobutadiene	15.198	225	13737	18.554	ug/l	99
95) Naphthalene	15.322	128	50618	19.045	ug/l	100
96) 1,2,3-Trichlorobenzene	15.516	180	23097	18.881	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101122\
Data File : VD074519.D
Acq On : 11 Oct 2022 15:17
Operator : VA/SY
Sample : VSTDICC020
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022

Quant Time: Oct 12 03:09:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 12 03:06:42 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101122\
 Data File : VD074519.D
 Acq On : 11 Oct 2022 15:17
 Operator : VA/SY
 Sample : VSTDICC020
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1


Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICC020

Manual Integrations APPROVED

Reviewed By :Krupa Patel 10/12/2022
 Supervised By :Mahesh Dadoda 10/13/2022


Quant Time: Oct 12 03:09:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 12 03:06:42 2022
 Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020



Tgt Ion:168 Resp: 92708
Ion Ratio Lower Upper
168 100
99 52.2 40.9 61.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#2
Dichlorodifluoromethane
Concen: 17.396 ug/l
RT: 1.934 min Scan# 59
Delta R.T. 0.006 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17


Tgt Ion: 85 Resp: 15557
Ion Ratio Lower Upper
85 100
87 34.3 17.5 52.6





#3
Chloromethane
Concen: 19.940 ug/l
RT: 2.146 min Scan# 95
Delta R.T. 0.006 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020




Tgt Ion: 50 Resp: 22094
Ion Ratio Lower Upper
50 100
52 30.1 30.0 45.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022




#4
Vinyl Chloride
Concen: 18.724 ug/l
RT: 2.281 min Scan# 118
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17



Tgt Ion: 62 Resp: 21475
Ion Ratio Lower Upper
62 100
64 34.8 25.5 38.3









#5
Bromomethane
Concen: 21.489 ug/l
RT: 2.670 min Scan# 187
Delta R.T. -0.011 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion: 94 Resp: 16740
Ion Ratio Lower Upper
94 100
96 88.0 74.6 112.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#6
Chloroethane
Concen: 20.419 ug/l
RT: 2.828 min Scan# 211
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Tgt Ion: 64 Resp: 15360
Ion Ratio Lower Upper
64 100
66 31.9 25.8 38.8









#7
 Trichlorofluoromethane
 Concen: 18.693 ug/l
 RT: 3.170 min Scan# 201
 Delta R.T. 0.001 min
 Lab File: VD074519.D
 Acq: 11 Oct 2022 15:17

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC020

Tgt Ion:101 Resp: 3242
 Ion Ratio Lower Upper
 101 100
 103 66.9 49.8 74.8




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 10/12/2022
 Supervised By :Mahesh Dadoda 10/13/2022



#8
 Diethyl Ether
 Concen: 19.072 ug/l
 RT: 3.593 min Scan# 341
 Delta R.T. 0.000 min
 Lab File: VD074519.D
 Acq: 11 Oct 2022 15:17


Tgt Ion: 74 Resp: 10137
 Ion Ratio Lower Upper
 74 100
 45 79.9 37.9 113.6





#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.447 ug/l
RT: 3.958 min Scan# 403
Delta R.T. -0.005 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020




Tgt Ion:101 Resp: 22818
Ion Ratio Lower Upper
101 100
85 43.6 34.1 51.1
151 71.8 61.7 92.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022




#10
Methyl Iodide
Concen: 18.620 ug/l
RT: 4.158 min Scan# 437
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Tgt Ion:142 Resp: 23876
Ion Ratio Lower Upper
142 100
127 40.7 31.9 47.9
141 14.8 11.4 17.2





#11
Tert butyl alcohol
Concen: 130.694 ug/l m
RT: 5.022 min Scan# 50
Delta R.T. -0.030 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020




Tgt Ion: 59 Resp: 16269
Ion Ratio Lower Upper
59 100
57 0.0 4.4 6.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022




#12
1,1-Dichloroethene
Concen: 19.468 ug/l
RT: 3.934 min Scan# 399
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Tgt Ion: 96 Resp: 23220
Ion Ratio Lower Upper
96 100
61 140.4 113.9 170.9
98 62.2 51.8 77.8





#13
Acrolein
Concen: 117.834 ug/l
RT: 3.793 min Scan# 31
Delta R.T. -0.006 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020



Tgt Ion: 56 Resp: 4584
Ion Ratio Lower Upper
56 100
55 65.7 54.0 81.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#14
Allyl chloride
Concen: 19.218 ug/l
RT: 4.558 min Scan# 505
Delta R.T. 0.006 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Tgt Ion: 41 Resp: 25275
Ion Ratio Lower Upper
41 100
39 67.8 53.4 80.2
76 51.6 42.4 63.6









#15
Acrylonitrile
Concen: 97.362 ug/l
RT: 5.252 min Scan# 61
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion: 53 Resp: 2041
Ion Ratio Lower Upper
53 100
52 84.9 66.0 99.0
51 37.8 28.3 42.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#16
Acetone
Concen: 83.264 ug/l
RT: 4.022 min Scan# 414
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Tgt Ion: 43 Resp: 14825
Ion Ratio Lower Upper
43 100
58 37.9 30.2 45.2









#17
Carbon Disulfide
Concen: 19.974 ug/l
RT: 4.269 min Scan# 456
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Instrument :
MSVOA_D
Client Sample Id :
VSTDIC020

Tgt Ion: 76 Resp: 7045
Ion Ratio Lower Upper
76 100
78 8.6 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#18
Methyl Acetate
Concen: 20.949 ug/l
RT: 4.564 min Scan# 506
Delta R.T. 0.006 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Tgt Ion: 43 Resp: 11242
Ion Ratio Lower Upper
43 100
74 29.7 24.7 37.1





#19

Methyl tert-butyl Ether

Concen: 20.312 ug/l

RT: 5.311 min Scan# 61

Delta R.T. -0.005 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC020

Tgt Ion: 73 Resp: 4629

Ion Ratio Lower Upper

73 100





57 22.9 16.1 24.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/12/2022

Supervised By :Mahesh Dadoda 10/13/2022



#20

Methylene Chloride

Concen: 22.369 ug/l

RT: 4.799 min Scan# 546

Delta R.T. -0.000 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Tgt Ion: 84 Resp: 36860




Ion Ratio Lower Upper


84 100

49 101.6 68.0 102.0

51 31.6 23.2 34.8

86 63.4 48.0 72.0









#21
trans-1,2-Dichloroethene
Concen: 19.285 ug/l
RT: 5.305 min Scan# 611
Delta R.T. -0.005 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion: 96 Resp: 2525
Ion Ratio Lower Upper
96 100
61 118.4 95.7 143.5
98 62.3 54.8 82.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#22
Diisopropyl ether
Concen: 21.220 ug/l
RT: 6.222 min Scan# 788
Delta R.T. 0.006 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Tgt Ion: 45 Resp: 55784
Ion Ratio Lower Upper
45 100
43 47.1 37.8 56.6
87 32.0 26.6 40.0
59 14.5 9.9 14.9









#23
 Vinyl Acetate
 Concen: 92.989 ug/l
 RT: 6.158 min Scan# 777
 Delta R.T. 0.001 min
 Lab File: VD074519.D
 Acq: 11 Oct 2022 15:17

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC020

Tgt Ion: 43 Resp: 10872
 Ion Ratio Lower Upper
 43 100
 86 13.3 11.2 16.8




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 10/12/2022
 Supervised By :Mahesh Dadoda 10/13/2022



#24
 1,1-Dichloroethane
 Concen: 20.111 ug/l
 RT: 6.111 min Scan# 769
 Delta R.T. 0.006 min
 Lab File: VD074519.D
 Acq: 11 Oct 2022 15:17

Tgt Ion: 63 Resp: 40572
 Ion Ratio Lower Upper
 63 100
 98 10.2 4.6 13.8
 100 5.0 2.9 8.6





#25

2-Butanone

Concen: 94.079 ug/l

RT: 7.087 min Scan# 91

Delta R.T. 0.006 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC020

Tgt Ion: 43 Resp: 22479

Ion Ratio Lower Upper

43 100





72 31.2 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/12/2022

Supervised By :Mahesh Dadoda 10/13/2022



#26

2,2-Dichloropropane

Concen: 19.545 ug/l

RT: 7.075 min Scan# 933

Delta R.T. 0.000 min

Lab File: VD074519.D




Acq: 11 Oct 2022 15:17


Tgt Ion: 77 Resp: 34994

Ion Ratio Lower Upper

77 100

97 23.9 12.5 37.5





#27

cis-1,2-Dichloroethene

Concen: 20.962 ug/l

RT: 7.075 min Scan# 91

Delta R.T. 0.000 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC020

Tgt Ion: 96 Resp: 28909





Ion	Ratio	Lower	Upper
96	100		
61	121.5	0.0	242.4
98	61.8	0.0	130.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/12/2022

Supervised By :Mahesh Dadoda 10/13/2022



#28

Bromochloromethane

Concen: 23.014 ug/l

RT: 7.428 min Scan# 993




Delta R.T. 0.000 min


Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Tgt Ion: 49 Resp: 12208

Ion	Ratio	Lower	Upper
49	100		
129	1.0	0.0	5.0
130	90.4	84.1	126.1





#29

Tetrahydrofuran

Concen: 102.735 ug/l

RT: 7.446 min Scan# 996

Delta R.T. -0.000 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC020





Tgt Ion:	42	Resp:	1348
Ion Ratio	Lower	Upper	
42	100		
72	53.7	45.8	68.6
71	55.1	44.6	66.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/12/2022

Supervised By :Mahesh Dadoda 10/13/2022



#30

Chloroform

Concen: 20.754 ug/l




RT: 7.593 min Scan# 1021


Delta R.T. -0.000 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17


Tgt Ion:	83	Resp:	41330
Ion Ratio	Lower	Upper	
83	100		
85	59.7	52.0	78.0





#31
Cyclohexane
Concen: 19.097 ug/l
RT: 7.881 min Scan# 1070
Delta R.T. 0.006 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020




Tgt Ion: 56 Resp: 3218
Ion Ratio Lower Upper
56 100
69 42.4 29.2 43.8
84 111.3 83.4 125.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022




#32
1,1,1-Trichloroethane
Concen: 19.705 ug/l
RT: 7.793 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17



Tgt Ion: 97 Resp: 35099
Ion Ratio Lower Upper
97 100
99 64.6 53.3 79.9
61 39.9 30.2 45.4





#33

1,2-Dichloroethane-d4

Concen: 19.261 ug/l

RT: 8.234 min Scan# 1130

Delta R.T. 0.006 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC020

Tgt Ion: 65 Resp: 1285

Ion Ratio Lower Upper

65 100





67 54.0 0.0 110.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/12/2022

Supervised By :Mahesh Dadoda 10/13/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1222

Delta R.T. 0.000 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17




Tgt Ion:114 Resp: 161053


Ion Ratio Lower Upper

114 100

63 16.9 0.0 33.2

88 14.4 0.0 28.2





#35

Dibromofluoromethane

Concen: 22.424 ug/l

RT: 7.805 min Scan# 1057

Delta R.T. 0.001 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC020





Tgt Ion:	113	Resp:	1979
Ion Ratio	Lower	Upper	
113	100		
111	100.5	82.3	123.5
192	19.8	16.4	24.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/12/2022

Supervised By :Mahesh Dadoda 10/13/2022



#36

1,1-Dichloropropene

Concen: 19.035 ug/l




RT: 8.010 min Scan# 1092


Delta R.T. 0.006 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17


Tgt Ion:	75	Resp:	29657
Ion Ratio	Lower	Upper	
75	100		
110	40.4	19.7	59.0
77	31.0	25.4	38.2





#37
Ethyl Acetate
Concen: 20.400 ug/l
RT: 7.169 min Scan# 949
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020



Tgt Ion: 43 Resp: 1024
Ion Ratio Lower Upper
43 100
61 15.5 12.2 18.4
70 10.6 9.6 14.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#38
Carbon Tetrachloride
Concen: 19.057 ug/l
RT: 7.987 min Scan# 1088
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Tgt Ion:117 Resp: 27307
Ion Ratio Lower Upper
117 100
119 98.9 73.1 109.7
121 35.0 25.0 37.6





#39

Methylcyclohexane

Concen: 19.685 ug/l

RT: 9.275 min Scan# 11307

Delta R.T. 0.006 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC020





Tgt Ion:	83	Resp:	3792
Ion	Ratio	Lower	Upper
83	100		
55	61.5	49.0	73.4
98	49.8	41.7	62.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/12/2022

Supervised By :Mahesh Dadoda 10/13/2022



#40

Benzene

Concen: 19.634 ug/l




RT: 8.252 min Scan# 1133


Delta R.T. -0.000 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Tgt Ion:	78	Resp:	90228
Ion	Ratio	Lower	Upper
78	100		
77	24.7	20.1	30.1






#41
Methacrylonitrile
Concen: 17.712 ug/l
RT: 7.399 min Scan# 911
Delta R.T. -0.005 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC020






Tgt Ion: 41 Resp: 535
Ion Ratio Lower Upper
41 100
39 55.6 35.0 52.6
67 97.8 77.0 115.6
52 34.9 29.7 44.5

Manual Integrations


APPROVED

Reviewed By :Krupa Patel 10/12/2022



Supervised By :Mahesh Dadoda 10/13/2022




#42
1,2-Dichloroethane
Concen: 19.294 ug/l
RT: 8.322 min Scan# 1145
Delta R.T. -0.006 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Tgt Ion: 62 Resp: 20880
Ion Ratio Lower Upper
62 100
98 12.4 0.0 24.2





#43
Isopropyl Acetate
Concen: 19.596 ug/l
RT: 8.363 min Scan# 1152
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020



Tgt Ion: 43 Resp: 18220
Ion Ratio Lower Upper
43 100
61 24.0 30.9 46.3
87 17.7 14.3 21.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#44
Trichloroethene
Concen: 19.857 ug/l
RT: 9.028 min Scan# 1265
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Tgt Ion:130 Resp: 25842
Ion Ratio Lower Upper
130 100
95 88.2 0.0 185.4





#45

1,2-Dichloropropane

Concen: 19.941 ug/l

RT: 9.305 min Scan# 1312

Delta R.T. 0.001 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC020

Tgt Ion: 63 Resp: 2086

Ion Ratio Lower Upper

63 100





65 33.7 23.1 34.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/12/2022

Supervised By :Mahesh Dadoda 10/13/2022



#46

Dibromomethane

Concen: 19.967 ug/l

RT: 9.393 min Scan# 1327

Delta R.T. -0.006 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17




Tgt Ion: 93 Resp: 11634


Ion Ratio Lower Upper

93 100

95 83.5 68.2 102.2

174 101.5 87.4 131.0





#47

Bromodichloromethane

Concen: 20.090 ug/l

RT: 9.581 min Scan# 1359

Delta R.T. -0.006 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC020





Tgt Ion: 83	Resp: 29114
Ion Ratio	Lower Upper
83 100	
85 54.4	48.8 73.2
127 7.1	6.7 10.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/12/2022

Supervised By :Mahesh Dadoda 10/13/2022



#48

Methyl methacrylate

Concen: 20.102 ug/l




RT: 9.375 min Scan# 1324


Delta R.T. -0.006 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Tgt Ion: 41	Resp: 8995
Ion Ratio	Lower Upper
41 100	
69 108.9	93.0 139.4
39 54.1	44.6 66.8





#49

1,4-Dioxane

Concen: 452.160 ug/l

RT: 9.387 min Scan# 1326

Delta R.T. -0.000 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC020

Tgt Ion: 88 Resp: 2650





Ion	Ratio	Lower	Upper
88	100		
43	21.4	24.1	36.1
58	51.7	43.1	64.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/12/2022

Supervised By :Mahesh Dadoda 10/13/2022



#50

Toluene-d8

Concen: 16.377 ug/l

RT: 10.269 min Scan# 1476




Delta R.T. 0.000 min


Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Tgt Ion: 98 Resp: 47400


Ion	Ratio	Lower	Upper
98	100		
100	66.8	51.4	77.0





#51
4-Methyl-2-Pentanone
Concen: 102.177 ug/l
RT: 10.157 min Scan# 1457
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020




Tgt Ion: 43 Resp: 46875
Ion Ratio Lower Upper
43 100
58 43.8 37.8 56.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022




#52
Toluene
Concen: 19.984 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17



Tgt Ion: 92 Resp: 58397
Ion Ratio Lower Upper
92 100
91 177.7 139.0 208.4









#53
 t-1,3-Dichloropropene
 Concen: 19.882 ug/l
 RT: 10.552 min Scan# 11
 Delta R.T. 0.001 min
 Lab File: VD074519.D
 Acq: 11 Oct 2022 15:17

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC020

Tgt Ion: 75 Resp: 2552
 Ion Ratio Lower Upper
 75 100
 77 34.5 25.6 38.4




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 10/12/2022
 Supervised By :Mahesh Dadoda 10/13/2022



#54
 cis-1,3-Dichloropropene
 Concen: 19.750 ug/l
 RT: 10.016 min Scan# 1433
 Delta R.T. 0.000 min
 Lab File: VD074519.D
 Acq: 11 Oct 2022 15:17


Tgt Ion: 75 Resp: 31832
 Ion Ratio Lower Upper
 75 100
 77 31.5 24.7 37.1
 39 34.5 25.4 38.2





#55
1,1,2-Trichloroethane
Concen: 20.709 ug/l
RT: 10.734 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020



Tgt Ion:	97	Resp:	1626
Ion Ratio	100	Lower	Upper
83	75.9	64.8	97.2
85	61.5	44.8	67.2
99	58.2	49.4	74.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#56
Ethyl methacrylate
Concen: 20.520 ug/l
RT: 10.593 min Scan# 1531
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17


Tgt Ion:	69	Resp:	18348
Ion Ratio	100	Lower	Upper
41	51.5	38.6	58.0
39	26.5	19.9	29.9





#57
1,3-Dichloropropane
Concen: 21.211 ug/l
RT: 10.881 min Scan# 1580
Delta R.T. 0.006 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020



Tgt Ion: 76 Resp: 2751
Ion Ratio Lower Upper
76 100
78 31.6 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#58
2-Chloroethyl Vinyl ether
Concen: 75.624 ug/l
RT: 9.869 min Scan# 1408
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Tgt Ion: 63 Resp: 13962
Ion Ratio Lower Upper
63 100
106 32.7 26.5 39.7









#59
2-Hexanone
Concen: 103.458 ug/l
RT: 10.916 min Scan# 1586
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion: 43 Resp: 3183
Ion Ratio Lower Upper
43 100
58 63.2 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#60
Dibromochloromethane
Concen: 20.776 ug/l
RT: 11.075 min Scan# 1613
Delta R.T. 0.006 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17


Tgt Ion:129 Resp: 19695
Ion Ratio Lower Upper
129 100
127 76.6 39.1 117.2





#61
1,2-Dibromoethane
Concen: 20.446 ug/l
RT: 11.181 min Scan# 1631
Delta R.T. 0.006 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020




Tgt Ion:107 Resp: 1491
Ion Ratio Lower Upper
107 100
109 93.4 74.6 111.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022




#62
4-Bromofluorobenzene
Concen: 28.269 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Tgt Ion: 95 Resp: 29984
Ion Ratio Lower Upper
95 100
174 90.2 0.0 197.0
176 85.6 0.0 182.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1699
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020




Tgt Ion:117 Resp: 14690
Ion Ratio Lower Upper
117 100
82 50.8 41.6 62.4
119 32.8 25.2 37.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022




#64
Tetrachloroethene
Concen: 19.753 ug/l
RT: 10.810 min Scan# 1568
Delta R.T. 0.006 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Tgt Ion:164 Resp: 21499
Ion Ratio Lower Upper
164 100
166 115.9 107.8 161.8
129 89.3 77.4 116.2
131 80.9 69.5 104.3





#65
Chlorobenzene
Concen: 20.147 ug/l
RT: 11.604 min Scan# 1703
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020




Tgt Ion:112 Resp: 63730
Ion Ratio Lower Upper
112 100
114 32.8 25.1 37.7

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 19.572 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17



Tgt Ion:131 Resp: 20872
Ion Ratio Lower Upper
131 100
133 99.2 46.6 139.8
119 70.5 32.5 97.4









#67
Ethyl Benzene
Concen: 19.491 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion: 91 Resp: 10928
Ion Ratio Lower Upper
91 100
106 34.4 26.6 40.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#68
m/p-Xylenes
Concen: 39.706 ug/l
RT: 11.793 min Scan# 1735
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Tgt Ion:106 Resp: 88638
Ion Ratio Lower Upper
106 100
91 191.6 156.2 234.4









#69
o-Xylene
Concen: 19.954 ug/l
RT: 12.116 min Scan# 1790
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion:106 Resp: 41300
Ion Ratio Lower Upper
106 100
91 205.5 104.5 313.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#70
Styrene
Concen: 20.476 ug/l
RT: 12.134 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Tgt Ion:104 Resp: 71802
Ion Ratio Lower Upper
104 100
78 42.9 35.7 53.5
103 54.4 43.7 65.5









#71
Bromoform
Concen: 20.210 ug/l
RT: 12.293 min Scan# 1820
Delta R.T. 0.001 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion:173 Resp: 10775
Ion Ratio Lower Upper
173 100
175 48.5 23.1 69.3
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.510 min Scan# 2027
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Tgt Ion:152 Resp: 68631
Ion Ratio Lower Upper
152 100
115 55.3 26.8 80.4
150 160.3 0.0 348.2









#73
Isopropylbenzene
Concen: 19.984 ug/l
RT: 12.416 min Scan# 1841
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion:105 Resp: 108999
Ion Ratio Lower Upper
105 100
120 26.9 13.5 40.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#74
N-ethyl acetate
Concen: 19.990 ug/l
RT: 12.228 min Scan# 1809
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17


Tgt Ion: 43 Resp: 17064
Ion Ratio Lower Upper
43 100
70 52.5 43.8 65.6
55 29.0 22.0 33.0
61 29.2 23.9 35.9





#75
1,1,2,2-Tetrachloroethane
Concen: 20.217 ug/l
RT: 12.669 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020



Tgt Ion: 83 Resp: 17498
Ion Ratio Lower Upper
83 100
131 10.8 4.7 14.1
85 64.2 33.1 99.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#76
1,2,3-Trichloropropane
Concen: 24.108 ug/l m
RT: 12.716 min Scan# 1892
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17


Tgt Ion: 75 Resp: 12409
Ion Ratio Lower Upper
75 100
77 0.0 185.8 557.5#





#77
Bromobenzene
Concen: 19.816 ug/l
RT: 12.693 min Scan# 1888
Delta R.T. -0.005 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020




Tgt Ion:156 Resp: 2454
Ion Ratio Lower Upper
156 100
77 160.2 77.7 233.1
158 97.3 49.4 148.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022




#78
n-propylbenzene
Concen: 20.204 ug/l
RT: 12.757 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17



Tgt Ion: 91 Resp: 133711
Ion Ratio Lower Upper
91 100
120 23.2 11.9 35.7









#79
2-Chlorotoluene
Concen: 20.297 ug/l
RT: 12.845 min Scan# 1914
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion: 91 Resp: 71733
Ion Ratio Lower Upper
91 100
126 36.5 18.3 54.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#80
1,3,5-Trimethylbenzene
Concen: 19.919 ug/l
RT: 12.898 min Scan# 1923
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17


Tgt Ion:105 Resp: 88409
Ion Ratio Lower Upper
105 100
120 50.7 25.7 77.0





#81
trans-1,4-Dichloro-2-butene
Concen: 17.553 ug/l
RT: 12.463 min Scan# 1850
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020




Tgt Ion: 75 Resp: 418
Ion Ratio Lower Upper
75 100
53 84.8 63.8 95.6
89 48.2 38.2 57.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022




#82
4-Chlorotoluene
Concen: 20.218 ug/l
RT: 12.945 min Scan# 1931
Delta R.T. 0.005 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17



Tgt Ion: 91 Resp: 72484
Ion Ratio Lower Upper
91 100
126 35.9 18.8 56.4









#83
tert-Butylbenzene
Concen: 19.770 ug/l
RT: 13.163 min Scan# 1968
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020

Tgt Ion:119 Resp: 76708
Ion Ratio Lower Upper
119 100
91 61.7 30.0 90.0
134 27.1 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#84
1,2,4-Trimethylbenzene
Concen: 19.988 ug/l
RT: 13.204 min Scan# 1975
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17


Tgt Ion:105 Resp: 87833
Ion Ratio Lower Upper
105 100
120 47.2 22.9 68.7





#85
 sec-Butylbenzene
 Concen: 19.778 ug/l
 RT: 13.340 min Scan# 1998
 Delta R.T. 0.001 min
 Lab File: VD074519.D
 Acq: 11 Oct 2022 15:17




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC020




Tgt Ion:105 Resp: 116810
 Ion Ratio Lower Upper
 105 100
 134 20.6 10.0 30.0

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 10/12/2022
 Supervised By :Mahesh Dadoda 10/13/2022




#86
 p-Isopropyltoluene
 Concen: 19.559 ug/l
 RT: 13.457 min Scan# 2018
 Delta R.T. 0.000 min
 Lab File: VD074519.D
 Acq: 11 Oct 2022 15:17



Tgt Ion:119 Resp: 96438
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.5 40.4
 91 22.1 10.8 32.4





#87

1,3-Dichlorobenzene

Concen: 19.876 ug/l

RT: 13.451 min Scan# 2017

Delta R.T. 0.000 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC020





Tgt Ion:146 Resp: 50058
 Ion Ratio Lower Upper
 146 100
 111 37.7 18.3 54.8
 148 63.3 31.8 95.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/12/2022

Supervised By :Mahesh Dadoda 10/13/2022



#88

1,4-Dichlorobenzene

Concen: 19.737 ug/l




RT: 13.534 min Scan# 2031


Delta R.T. -0.000 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Tgt Ion:146 Resp: 49164
 Ion Ratio Lower Upper
 146 100
 111 37.0 17.7 53.1
 148 63.8 32.3 96.9









#89
n-Butylbenzene
Concen: 19.895 ug/l
RT: 13.781 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC020

Tgt Ion: 91 Resp: 9165
Ion Ratio Lower Upper
91 100
92 51.8 26.9 80.7
134 26.7 13.4 40.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#90
Hexachloroethane
Concen: 19.237 ug/l
RT: 14.045 min Scan# 2118
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17

Tgt Ion:117 Resp: 16834
Ion Ratio Lower Upper
117 100
201 69.4 38.0 114.0





#91

1,2-Dichlorobenzene

Concen: 19.778 ug/l

RT: 13.828 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC020

Tgt Ion:146 Resp: 4260





Ion	Ratio	Lower	Upper
146	100		
111	39.6	18.9	56.5
148	62.9	31.9	95.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/12/2022

Supervised By :Mahesh Dadoda 10/13/2022



#92

1,2-Dibromo-3-Chloropropane

Concen: 19.758 ug/l

RT: 14.440 min Scan# 2185




Delta R.T. 0.001 min


Lab File: VD074519.D

Acq: 11 Oct 2022 15:17

Tgt Ion: 75 Resp: 2406


Ion	Ratio	Lower	Upper
75	100		
155	87.3	51.2	153.8
157	118.4	68.2	204.5





#93
1,2,4-Trichlorobenzene
Concen: 19.121 ug/l
RT: 15.092 min Scan# 2296
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020




Tgt Ion:180 Resp: 2710
Ion Ratio Lower Upper
180 100
182 97.1 46.9 140.6
145 29.6 14.1 42.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022




#94
Hexachlorobutadiene
Concen: 18.554 ug/l
RT: 15.198 min Scan# 2314
Delta R.T. 0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Tgt Ion:225 Resp: 13737
Ion Ratio Lower Upper
225 100
223 65.3 31.9 95.5
227 64.5 32.6 97.8





#95
Naphthalene
Concen: 19.045 ug/l
RT: 15.322 min Scan# 2335
Delta R.T. -0.006 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC020




Tgt Ion:128 Resp: 50618
Ion Ratio Lower Upper
128 100
127 13.4 10.6 16.0
129 10.7 8.3 12.5

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 10/12/2022
Supervised By :Mahesh Dadoda 10/13/2022



#96
1,2,3-Trichlorobenzene
Concen: 18.881 ug/l
RT: 15.516 min Scan# 2368
Delta R.T. -0.000 min
Lab File: VD074519.D
Acq: 11 Oct 2022 15:17



Tgt Ion:180 Resp: 23097
Ion Ratio Lower Upper
180 100
182 96.5 46.7 140.1
145 32.6 14.9 44.7

