



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101122\
 Data File : VD074510.D
 Acq On : 11 Oct 2022 10:00
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
 Title : SW846 8260
 Last Update : Wed Oct 12 13:14:00 2022



AutoFind: Scans 1866, 1867, 1868; Background Corrected with Scan 1858

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	10115	PASS
75	95	30	60	50.6	28357	PASS
95	95	100	100	100.0	56048	PASS
96	95	5	9	7.2	4045	PASS
173	174	0.00	2	1.6	845	PASS
174	95	50	100	93.9	52629	PASS
175	174	5	9	7.4	3872	PASS
176	174	95	101	95.0	49979	PASS
177	176	5	9	6.8	3409	PASS