

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102522\
 Data File : VD074625.D
 Acq On : 25 Oct 2022 11:41
 Operator : VA/SY
 Sample : VD1025SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1025SBL01

Quant Time: Oct 27 01:20:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	91342	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	153749	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	140790	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	64143	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	34224	49.075	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.160%
35) Dibromofluoromethane	7.804	113	42849	50.309	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.620%
50) Toluene-d8	10.269	98	147206	47.836	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.680%
62) 4-Bromofluorobenzene	12.569	95	47581	53.897	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102522\
Data File : VD074625.D
Acq On : 25 Oct 2022 11:41
Operator : VA/SY
Sample : VD1025SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1025SBL01

Quant Time: Oct 27 01:20:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:41:20 2022
Response via : Initial Calibration

