

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102522\
 Data File : VD074630.D
 Acq On : 25 Oct 2022 14:01
 Operator : VA/SY
 Sample : N5218-11RE
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-2RE

Quant Time: Oct 27 01:21:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	1664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	3018	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	2589	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	1334	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	1369	107.758	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 215.520%#	
35) Dibromofluoromethane	7.810	113	1259	75.865	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 151.720%#	
50) Toluene-d8	10.269	98	3830	63.404	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 126.800%	
62) 4-Bromofluorobenzene	12.569	95	1066	61.800	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 123.600%	
Target Compounds						
20) Methylene Chloride	4.822	84	56	Below Cal	Qvalue #	18

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102522\
Data File : VD074630.D
Acq On : 25 Oct 2022 14:01
Operator : VA/SY
Sample : N5218-11RE
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-2RE

Quant Time: Oct 27 01:21:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:41:20 2022
Response via : Initial Calibration

