

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102522\
 Data File : VD074625.D
 Acq On : 25 Oct 2022 11:41
 Operator : VA/SY
 Sample : VD1025SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1025SBL01

Quant Time: Oct 27 01:20:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	91342	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	153749	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	140790	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	64143	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	34224	49.075	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.160%
35) Dibromofluoromethane	7.804	113	42849	50.309	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.620%
50) Toluene-d8	10.269	98	147206	47.836	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.680%
62) 4-Bromofluorobenzene	12.569	95	47581	53.897	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.800%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102522\
Data File : VD074625.D
Acq On : 25 Oct 2022 11:41
Operator : VA/SY
Sample : VD1025SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1025SBL01

Quant Time: Oct 27 01:20:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:41:20 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1129
Delta R.T. 0.000 min
Lab File: VD074625.D
Acq: 25 Oct 2022 11:41




Instrument :
MSVOA_D
ClientSampleId :
VD1025SBL01


Tgt Ion:168 Resp: 91342
Ion Ratio Lower Upper
168 100
99 47.9 40.9 61.3



#33
1,2-Dichloroethane-d4
Concen: 49.075 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. -0.000 min
Lab File: VD074625.D
Acq: 25 Oct 2022 11:41

Tgt Ion: 65 Resp: 34224
Ion Ratio Lower Upper
65 100
67 54.7 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD074625.D

Acq: 25 Oct 2022 11:41

Instrument :

MSVOA_D

ClientSampleId :

VD1025SBL01





Tgt Ion:114 Resp: 153749

Ion Ratio Lower Upper

114 100

63 17.1 0.0 33.2

88 13.7 0.0 28.2



#35

Dibromofluoromethane

Concen: 50.309 ug/l

RT: 7.804 min Scan# 1057

Delta R.T. 0.000 min

Lab File: VD074625.D

Acq: 25 Oct 2022 11:41




Tgt Ion:113 Resp: 42849


Ion Ratio Lower Upper

113 100

111 102.3 82.3 123.5

192 19.1 16.4 24.6









#50
Toluene-d8
Concen: 47.836 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD074625.D
Acq: 25 Oct 2022 11:41




Instrument :
MSVOA_D
ClientSampleId :
VD1025SBL01


Tgt Ion: 98 Resp: 147206
Ion Ratio Lower Upper
98 100
100 64.0 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 53.897 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VD074625.D
Acq: 25 Oct 2022 11:41

Tgt Ion: 95 Resp: 47581
Ion Ratio Lower Upper
95 100
174 93.2 0.0 197.0
176 92.1 0.0 182.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD074625.D
Acq: 25 Oct 2022 11:41

Instrument :
MSVOA_D
ClientSampleId :
VD1025SBL01

Tgt Ion:117 Resp: 140790
Ion Ratio Lower Upper
117 100
82 48.5 41.6 62.4
119 31.7 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. 0.006 min
Lab File: VD074625.D
Acq: 25 Oct 2022 11:41

Tgt Ion:152 Resp: 64143
Ion Ratio Lower Upper
152 100
115 55.5 26.8 80.4
150 159.0 0.0 348.2

