

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD102522\  
 Data File : VD074632.D  
 Acq On : 25 Oct 2022 14:57  
 Operator : VA/SY  
 Sample : N5267-04  
 Misc : 5.02G/5.00ml/MSVOA\_D/SOIL  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 TP2-1024

Quant Time: Oct 27 01:21:46 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D101922S.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 20 06:41:20 2022  
 Response via : Initial Calibration

| Compound                   | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| -----                      |       |      |          |       |       |          |
| Internal Standards         |       |      |          |       |       |          |
| 1) Pentafluorobenzene      | 0.000 | 168  | 0        | 0.000 | ug/l  | -7.88    |
| 34) 1,4-Difluorobenzene    | 0.000 | 114  | 0        | 0.000 | ug/l  | -8.78    |
| 63) Chlorobenzene-d5       | 0.000 | 117  | 0        | 0.000 | ug/l  | -11.58   |
| 72) 1,4-Dichlorobenzene-d4 | 0.000 | 152  | 0        | 0.000 | ug/l  | -13.51   |

|                             |        |       |          |          |      |         |
|-----------------------------|--------|-------|----------|----------|------|---------|
| System Monitoring Compounds |        |       |          |          |      |         |
| 33) 1,2-Dichloroethane-d4   | 0.000  | 65    | 0        | 0.000    | ug/l |         |
| Spiked Amount               | 50.000 | Range | 50 - 163 | Recovery | =    | 0.000%# |
| 35) Dibromofluoromethane    | 0.000  | 113   | 0        | 0.000    | ug/l |         |
| Spiked Amount               | 50.000 | Range | 54 - 147 | Recovery | =    | 0.000%# |
| 50) Toluene-d8              | 0.000  | 98    | 0        | 0.000    | ug/l |         |
| Spiked Amount               | 50.000 | Range | 49 - 140 | Recovery | =    | 0.000%# |
| 62) 4-Bromofluorobenzene    | 0.000  | 95    | 0        | 0.000    | ug/l |         |
| Spiked Amount               | 50.000 | Range | 25 - 144 | Recovery | =    | 0.000%# |


| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD102522\  
Data File : VD074632.D  
Acq On : 25 Oct 2022 14:57  
Operator : VA/SY  
Sample : N5267-04  
Misc : 5.02G/5.00ml/MSVOA\_D/SOIL  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_D  
ClientSampleId :  
TP2-1024

Quant Time: Oct 27 01:21:46 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D101922S.M  
Quant Title : SW846 8260  
QLast Update : Thu Oct 20 06:41:20 2022  
Response via : Initial Calibration







#1  
Pentafluorobenzene  
Concen: 0.000 ug/l  
Expected RT: 7.88 min  
  
Lab File: VD074632.D  
Acq: 25 Oct 2022 14:57


Instrument :  
MSVOA\_D  
ClientSampleId :  
TP2-1024


Tgt Ion: 168  
Sig Exp Ratio  
168 100  
99 51.1



#33  
1,2-Dichloroethane-d4  
Concen: 0.000 ug/l  
Expected RT: 8.23 min  
  
Lab File: VD074632.D  
Acq: 25 Oct 2022 14:57

Tgt Ion: 65  
Sig Exp Ratio  
65 100  
67 55.1







#34  
 1,4-Difluorobenzene  
 Concen: 0.000 ug/l  
 Expected RT: 8.78 min  
 Lab File: VD074632.D  
 Acq: 25 Oct 2022 14:57

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 TP2-1024

Tgt Ion: 114  


| Sig | Exp Ratio |
|-----|-----------|
| 114 | 100       |
| 63  | 16.6      |
| 88  | 14.1      |




#35  
 Dibromofluoromethane  
 Concen: 0.000 ug/l  
 Expected RT: 7.80 min  
 Lab File: VD074632.D  
 Acq: 25 Oct 2022 14:57

Tgt Ion: 113  

| Sig | Exp Ratio |
|-----|-----------|
| 113 | 100       |
| 111 | 102.9     |
| 192 | 20.5      |







#50  
Toluene-d8  
Concen: 0.000 ug/l  
Expected RT: 10.27 min  
  
Lab File: VD074632.D  
Acq: 25 Oct 2022 14:57


Instrument :  
MSVOA\_D  
ClientSampleId :  
TP2-1024


Tgt Ion: 98  
Sig Exp Ratio  
98 100  
100 64.2



#62  
4-Bromofluorobenzene  
Concen: 0.000 ug/l  
Expected RT: 12.57 min  
  
Lab File: VD074632.D  
Acq: 25 Oct 2022 14:57

Tgt Ion: 95  
Sig Exp Ratio  
95 100  
174 98.5  
176 91.3







#63  
 Chlorobenzene-d5  
 Concen: 0.000 ug/l  
 Expected RT: 11.58 min  
 Lab File: VD074632.D  
 Acq: 25 Oct 2022 14:57

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 TP2-1024

Tgt Ion: 117  

| Sig | Exp Ratio |
|-----|-----------|
| 117 | 100       |
| 82  | 52.0      |
| 119 | 31.5      |



#72  
 1,4-Dichlorobenzene-d4  
 Concen: 0.000 ug/l  
 Expected RT: 13.51 min  
 Lab File: VD074632.D  
 Acq: 25 Oct 2022 14:57

Tgt Ion: 152  

| Sig | Exp Ratio |
|-----|-----------|
| 152 | 100       |
| 115 | 53.6      |
| 150 | 174.1     |

