



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102522\
 Data File : VD074623.D
 Acq On : 25 Oct 2022 09:18
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Title : SW846 8260
 Last Update : Thu Oct 20 06:41:20 2022



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1858

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.5	9452	PASS
75	95	30	60	49.7	30307	PASS
95	95	100	100	100.0	61000	PASS
96	95	5	9	6.4	3903	PASS
173	174	0.00	2	1.5	908	PASS
174	95	50	100	97.8	59680	PASS
175	174	5	9	7.7	4590	PASS
176	174	95	101	95.7	57117	PASS
177	176	5	9	6.3	3621	PASS