

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102822\
 Data File : VD074670.D
 Acq On : 28 Oct 2022 14:35
 Operator : VA/SY
 Sample : N5282-01
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SUMP-STONE

Quant Time: Oct 29 02:14:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 00:18:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	91176	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	167408	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	133249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	52645	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	58205	75.636	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	151.280%
35) Dibromofluoromethane	7.805	113	63366	61.988	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	123.980%
50) Toluene-d8	10.269	98	213054	54.196	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.400%
62) 4-Bromofluorobenzene	12.575	95	51602	47.795	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102822\
Data File : VD074670.D
Acq On : 28 Oct 2022 14:35
Operator : VA/SY
Sample : N5282-01
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SUMP-STONE

Quant Time: Oct 29 02:14:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 00:18:17 2022
Response via : Initial Calibration

