

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102822\
 Data File : VD074675.D
 Acq On : 28 Oct 2022 16:52
 Operator : VA/SY
 Sample : N5313-23
 Misc : 10.78G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TOTAL-VOC-GRAB-3-SE-QUARTER

Quant Time: Oct 29 02:15:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 00:18:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	16734	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	31132	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	28086	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	10509	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	16678	118.084	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	236.160%#
35) Dibromofluoromethane	7.810	113	15021	79.017	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	158.040%#
50) Toluene-d8	10.269	98	44508	60.881	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	121.760%
62) 4-Bromofluorobenzene	12.575	95	11013	54.852	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102822\
Data File : VD074675.D
Acq On : 28 Oct 2022 16:52
Operator : VA/SY
Sample : N5313-23
Misc : 10.78G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TOTAL-VOC-GRAB-3-SE-QUARTER

Quant Time: Oct 29 02:15:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 00:18:17 2022
Response via : Initial Calibration

