

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102822\
 Data File : VD074678.D
 Acq On : 28 Oct 2022 18:15
 Operator : VA/SY
 Sample : N5267-13
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP5-1024

Quant Time: Oct 29 02:16:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 00:18:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	96491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	173916	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	155125	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	68825	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	61812	75.899	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	151.800%
35) Dibromofluoromethane	7.804	113	72321	68.101	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	136.200%
50) Toluene-d8	10.269	98	244335	59.827	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	119.660%
62) 4-Bromofluorobenzene	12.575	95	71855	64.063	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	128.120%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102822\
Data File : VD074678.D
Acq On : 28 Oct 2022 18:15
Operator : VA/SY
Sample : N5267-13
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP5-1024

Quant Time: Oct 29 02:16:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 00:18:17 2022
Response via : Initial Calibration

