

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102822\
 Data File : VD074680.D
 Acq On : 28 Oct 2022 19:10
 Operator : VA/SY
 Sample : N5284-04
 Misc : 6.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-3-WC

Quant Time: Oct 29 02:16:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 00:18:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	157402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	273446	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	211649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	81589	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	80639	60.699	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	121.400%	
35) Dibromofluoromethane	7.805	113	99149	59.380	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	118.760%	
50) Toluene-d8	10.269	98	358046	55.760	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	111.520%	
62) 4-Bromofluorobenzene	12.569	95	84775	48.071	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	96.140%	
Target Compounds						Qvalue
16) Acetone	4.022	43	10610	35.360	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102822\
Data File : VD074680.D
Acq On : 28 Oct 2022 19:10
Operator : VA/SY
Sample : N5284-04
Misc : 6.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-3-WC

Quant Time: Oct 29 02:16:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 00:18:17 2022
Response via : Initial Calibration

