

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102822\
 Data File : VD074669.D
 Acq On : 28 Oct 2022 14:07
 Operator : VA/SY
 Sample : N5213-06
 Misc : 11.36G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TOTAL-VOC-GRAB-5-SW-QUARTER

Quant Time: Oct 29 02:14:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 00:18:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	181052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	307590	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	239123	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	102956	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	85405	55.889	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.780%
35) Dibromofluoromethane	7.805	113	106586	56.748	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.500%
50) Toluene-d8	10.269	98	405654	56.161	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.320%
62) 4-Bromofluorobenzene	12.575	95	102375	51.607	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.220%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102822\
Data File : VD074669.D
Acq On : 28 Oct 2022 14:07
Operator : VA/SY
Sample : N5213-06
Misc : 11.36G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TOTAL-VOC-GRAB-5-SW-QUARTER

Quant Time: Oct 29 02:14:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 00:18:17 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD074669.D
Acq: 28 Oct 2022 14:07




Instrument :
MSVOA_D
ClientSampleId :
TOTAL-VOC-GRAB-5-SW-QUARTER


Tgt Ion:168 Resp: 181052
Ion Ratio Lower Upper
168 100
99 51.6 40.9 61.3



#33
1,2-Dichloroethane-d4
Concen: 55.889 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. -0.000 min
Lab File: VD074669.D
Acq: 28 Oct 2022 14:07

Tgt Ion: 65 Resp: 85405
Ion Ratio Lower Upper
65 100
67 54.6 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD074669.D

Acq: 28 Oct 2022 14:07

Instrument :

MSVOA_D

ClientSampleId :

TOTAL-VOC-GRAB-5-SW-QUARTER





Tgt Ion:114 Resp: 307590

Ion Ratio Lower Upper

114 100

63 16.3 0.0 33.2

88 13.3 0.0 28.2



#35

Dibromofluoromethane

Concen: 56.748 ug/l

RT: 7.805 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VD074669.D

Acq: 28 Oct 2022 14:07




Tgt Ion:113 Resp: 106586


Ion Ratio Lower Upper

113 100

111 103.2 82.3 123.5

192 19.8 16.4 24.6









#50
Toluene-d8
Concen: 56.161 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD074669.D
Acq: 28 Oct 2022 14:07




Instrument :
MSVOA_D
ClientSampleId :
TOTAL-VOC-GRAB-5-SW-QUARTER


Tgt Ion: 98 Resp: 405654
Ion Ratio Lower Upper
98 100
100 64.3 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 51.607 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VD074669.D
Acq: 28 Oct 2022 14:07

Tgt Ion: 95 Resp: 102375
Ion Ratio Lower Upper
95 100
174 92.9 0.0 197.0
176 89.0 0.0 182.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 117
Delta R.T. -0.000 min
Lab File: VD074669.D
Acq: 28 Oct 2022 14:07

Instrument :
MSVOA_D
ClientSampleId :
TOTAL-VOC-GRAB-5-SW-QUARTER

Tgt Ion:117 Resp: 239123
Ion Ratio Lower Upper
117 100
82 49.1 41.6 62.4
119 30.6 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.000 min
Lab File: VD074669.D
Acq: 28 Oct 2022 14:07

Tgt Ion:152 Resp: 102956
Ion Ratio Lower Upper
152 100
115 53.9 26.8 80.4
150 154.3 0.0 348.2

