



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110222\
 Data File : VD074698.D
 Acq On : 02 Nov 2022 11:25
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Title : SW846 8260
 Last Update : Thu Nov 03 02:30:45 2022



AutoFind: Scans 1866, 1867, 1868; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	8279	PASS
75	95	30	60	47.4	24536	PASS
95	95	100	100	100.0	51800	PASS
96	95	5	9	6.9	3596	PASS
173	174	0.00	2	1.1	534	PASS
174	95	50	100	94.3	48845	PASS
175	174	5	9	7.3	3586	PASS
176	174	95	101	97.9	47813	PASS
177	176	5	9	7.0	3331	PASS