

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
 Data File : VD074758.D
 Acq On : 07 Nov 2022 13:45
 Operator : VA/SY
 Sample : N5418-16RE
 Misc : 4.46G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-ORE

Quant Time: Nov 08 00:21:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	3040	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	9825	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	12652	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	6283	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	4094	170.035	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 340.060%#	
35) Dibromofluoromethane	7.805	113	5310	91.365	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 182.720%#	
50) Toluene-d8	10.269	98	12756	60.170	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 120.340%	
62) 4-Bromofluorobenzene	12.569	95	4838	79.038	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 158.080%#	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
Data File : VD074758.D
Acq On : 07 Nov 2022 13:45
Operator : VA/SY
Sample : N5418-16RE
Misc : 4.46G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-ORE

Quant Time: Nov 08 00:21:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

