

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
 Data File : VD074765.D
 Acq On : 07 Nov 2022 17:01
 Operator : VA/SY
 Sample : N5494-11
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-24

Quant Time: Nov 08 00:22:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	6694	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	12689	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	11855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	5767	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	7714	145.498	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	291.000%#
35) Dibromofluoromethane	7.805	113	7030	93.658	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	187.320%#
50) Toluene-d8	10.269	98	15759	57.557	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.120%
62) 4-Bromofluorobenzene	12.575	95	5552	70.230	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	140.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
Data File : VD074765.D
Acq On : 07 Nov 2022 17:01
Operator : VA/SY
Sample : N5494-11
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-24

Quant Time: Nov 08 00:22:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

