

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
 Data File : VD074769.D
 Acq On : 07 Nov 2022 18:54
 Operator : VA/SY
 Sample : N5494-19
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-21

Quant Time: Nov 08 00:22:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	62333	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	100171	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	97402	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	36118	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	44682	90.506	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 181.020%#	
35) Dibromofluoromethane	7.804	113	54330	91.688	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 183.380%#	
50) Toluene-d8	10.269	98	131399	60.792	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 121.580%	
62) 4-Bromofluorobenzene	12.569	95	39431	63.182	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 126.360%	
Target Compounds						
					Qvalue	
16) Acetone	4.022	43	2794	27.120	ug/l #	48
20) Methylene Chloride	4.793	84	15471	8.351	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
Data File : VD074769.D
Acq On : 07 Nov 2022 18:54
Operator : VA/SY
Sample : N5494-19
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-21

Quant Time: Nov 08 00:22:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

