

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
 Data File : VD074777.D
 Acq On : 08 Nov 2022 00:01
 Operator : VA/SY
 Sample : N5415-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R-14

Quant Time: Nov 08 02:53:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	3162	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	6305	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	6841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	3412	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	4666	186.315	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	372.620%#
35) Dibromofluoromethane	7.805	113	3421	91.724	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	183.440%#
50) Toluene-d8	10.269	98	7278	53.496	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.000%
62) 4-Bromofluorobenzene	12.575	95	3287	83.679	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	167.360%#
Target Compounds						
						Qvalue
3) Chloromethane	2.140	50	652	15.733	ug/l	# 86
10) Methyl Iodide	4.169	142	596	17.735	ug/l	# 43

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
Data File : VD074777.D
Acq On : 08 Nov 2022 00:01
Operator : VA/SY
Sample : N5415-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R-14

Quant Time: Nov 08 02:53:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

