

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
 Data File : VD074778.D
 Acq On : 08 Nov 2022 00:29
 Operator : VA/SY
 Sample : N5415-02
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R-15

Quant Time: Nov 08 02:58:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	7671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	16157	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	15409	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	7469	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	9309	153.220	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	306.440%#	
35) Dibromofluoromethane	7.811	113	8555	89.511	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	179.020%#	
50) Toluene-d8	10.269	98	17703	50.779	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	101.560%	
62) 4-Bromofluorobenzene	12.569	95	6634	65.905	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	131.800%	
Target Compounds						
20) Methylene Chloride	4.799	84	3388	26.026	ug/l	Qvalue # 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
Data File : VD074778.D
Acq On : 08 Nov 2022 00:29
Operator : VA/SY
Sample : N5415-02
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R-15

Quant Time: Nov 08 02:58:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

