

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
 Data File : VD074784.D
 Acq On : 08 Nov 2022 03:15
 Operator : VA/SY
 Sample : N5415-08
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R-21

Quant Time: Nov 08 05:09:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	1149	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	2479	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	2304	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	708	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	1014	111.425	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	222.840%#
35) Dibromofluoromethane	7.799	113	1365	93.083	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	186.160%#
50) Toluene-d8	10.263	98	3006	56.196	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.400%
62) 4-Bromofluorobenzene	12.569	95	936	60.604	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	121.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
Data File : VD074784.D
Acq On : 08 Nov 2022 03:15
Operator : VA/SY
Sample : N5415-08
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R-21

Quant Time: Nov 08 05:09:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

