

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
 Data File : VD074786.D
 Acq On : 08 Nov 2022 04:11
 Operator : VA/SY
 Sample : N5415-10
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R-23

Quant Time: Nov 08 05:15:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	111557	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	198044	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	180243	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	75103	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	61404	69.497	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	139.000%
35) Dibromofluoromethane	7.810	113	82194	70.161	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	140.320%
50) Toluene-d8	10.269	98	251545	58.864	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	117.720%
62) 4-Bromofluorobenzene	12.574	95	70113	56.825	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.640%
Target Compounds						
					Qvalue	
16) Acetone	4.016	43	2639	14.313	ug/l	92
20) Methylene Chloride	4.798	84	37387	16.294	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
Data File : VD074786.D
Acq On : 08 Nov 2022 04:11
Operator : VA/SY
Sample : N5415-10
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R-23

Quant Time: Nov 08 05:15:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

