

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
 Data File : VD074787.D
 Acq On : 08 Nov 2022 04:38
 Operator : VA/SY
 Sample : N5415-11
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R-24

Quant Time: Nov 08 05:20:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	3491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	6482	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	6909	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	3768	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	4679	169.226	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	338.460%#	
35) Dibromofluoromethane	7.804	113	3991	104.085	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	208.180%#	
50) Toluene-d8	10.269	98	8499	60.765	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	121.540%	
62) 4-Bromofluorobenzene	12.569	95	3258	80.676	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	161.360%#	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.804	84	2170	42.465	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
Data File : VD074787.D
Acq On : 08 Nov 2022 04:38
Operator : VA/SY
Sample : N5415-11
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R-24

Quant Time: Nov 08 05:20:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

