

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
 Data File : VD074788.D
 Acq On : 08 Nov 2022 05:06
 Operator : VA/SY
 Sample : N5415-12
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R-25

Quant Time: Nov 08 05:25:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	6765	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	11905	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	13667	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	6194	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	8455	157.801	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 315.600%#	
35) Dibromofluoromethane	7.804	113	8102	115.048	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 230.100%#	
50) Toluene-d8	10.269	98	15713	61.168	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 122.340%	
62) 4-Bromofluorobenzene	12.569	95	6003	80.936	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 161.880%#	
Target Compounds						Qvalue
20) Methylene Chloride	4.799	84	4327	44.111	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
Data File : VD074788.D
Acq On : 08 Nov 2022 05:06
Operator : VA/SY
Sample : N5415-12
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R-25

Quant Time: Nov 08 05:25:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

