

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
 Data File : VD074792.D
 Acq On : 08 Nov 2022 06:56
 Operator : VA/SY
 Sample : N5415-16
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R-29

Quant Time: Nov 08 07:57:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	106224	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	192856	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	186728	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	82858	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	58591	69.642	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 139.280%	
35) Dibromofluoromethane	7.805	113	74901	65.655	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 131.320%	
50) Toluene-d8	10.269	98	253192	60.843	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 121.680%	
62) 4-Bromofluorobenzene	12.575	95	75257	62.635	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 125.260%	
Target Compounds						
20) Methylene Chloride	4.805	84	50349	28.979	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
Data File : VD074792.D
Acq On : 08 Nov 2022 06:56
Operator : VA/SY
Sample : N5415-16
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R-29

Quant Time: Nov 08 07:57:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

