

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
 Data File : VD074751.D
 Acq On : 07 Nov 2022 09:53
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 07 11:44:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.411	0.344	16.3	94	0.00
3 P	Chloromethane	0.655	0.595	9.2	98	0.00
4 C	Vinyl Chloride	0.719	0.677	5.8#	97	0.00
5 T	Bromomethane	0.545	0.492	9.7	104	0.00
6 T	Chloroethane	0.498	0.485	2.6	101	0.00
7 T	Trichlorofluoromethane	0.835	0.787	5.7	97	0.00
8 T	Diethyl Ether	0.244	0.237	2.9	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.530	0.503	5.1	100	0.00
10 T	Methyl Iodide	0.571	0.543	4.9	86	0.00
11 T	Tert butyl alcohol	0.070	0.030	57.1#	82	0.01
12 CM	1,1-Dichloroethene	0.539	0.510	5.4#	95	0.00
13 T	Acrolein	0.027	0.027	0.0	109	0.00
14 T	Allyl chloride	0.570	0.581	-1.9	103	0.00
15 T	Acrylonitrile	0.096	0.097	-1.0	103	0.00
16 T	Acetone	0.083	0.097	-16.9	118	0.00
17 T	Carbon Disulfide	1.624	1.514	6.8	94	0.00
18 T	Methyl Acetate	0.250	0.227	9.2	104	0.00
19 T	Methyl tert-butyl Ether	1.024	1.060	-3.5	104	0.00
20 T	Methylene Chloride	1.061	0.640	39.7#	95	0.00
21 T	trans-1,2-Dichloroethene	0.592	0.578	2.4	99	0.00
22 T	Diisopropyl ether	1.176	1.257	-6.9	104	0.00
23 T	Vinyl Acetate	0.500	0.598	-19.6	121	0.00
24 P	1,1-Dichloroethane	0.900	0.887	1.4	102	0.00
25 T	2-Butanone	0.114	0.119	-4.4	108	0.00
26 T	2,2-Dichloropropane	0.828	0.812	1.9	103	0.00
27 T	cis-1,2-Dichloroethene	0.653	0.638	2.3	99	0.00
28 T	Bromochloromethane	0.284	0.303	-6.7	111	0.00
29 T	Tetrahydrofuran	0.065	0.068	-4.6	103	0.00
30 C	Chloroform	0.961	0.957	0.4#	102	0.00
31 T	Cyclohexane	0.802	0.741	7.6	96	0.00
32 T	1,1,1-Trichloroethane	0.849	0.831	2.1	100	0.00
33 S	1,2-Dichloroethane-d4	0.396	0.399	-0.8	110	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.296	0.298	-0.7	111	0.00
36 T	1,1-Dichloropropene	0.449	0.443	1.3	99	0.00
37 T	Ethyl Acetate	0.150	0.152	-1.3	105	0.00
38 T	Carbon Tetrachloride	0.425	0.418	1.6	100	0.00
39 T	Methylcyclohexane	0.546	0.547	-0.2	98	0.00
40 TM	Benzene	1.327	1.327	0.0	102	0.00
41 T	Methacrylonitrile	0.079	0.075	5.1	93	0.00
42 TM	1,2-Dichloroethane	0.305	0.314	-3.0	105	0.00
43 T	Isopropyl Acetate	0.279	0.282	-1.1	104	0.00
44 TM	Trichloroethene	0.378	0.374	1.1	101	0.00
45 C	1,2-Dichloropropane	0.293	0.307	-4.8#	104	0.00
46 T	Dibromomethane	0.170	0.183	-7.6	107	0.00
47 T	Bromodichloromethane	0.409	0.436	-6.6	105	0.00
48 T	Methyl methacrylate	0.124	0.133	-7.3	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
 Data File : VD074751.D
 Acq On : 07 Nov 2022 09:53
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 07 11:44:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	100	0.00
50 S	Toluene-d8	1.079	1.120	-3.8	121	0.00
51 T	4-Methyl-2-Pentanone	0.135	0.148	-9.6	109	0.00
52 CM	Toluene	0.812	0.871	-7.3#	114	0.00
53 T	t-1,3-Dichloropropene	0.362	0.408	-12.7	121	0.00
54 T	cis-1,3-Dichloropropene	0.454	0.490	-7.9	105	0.00
55 T	1,1,2-Trichloroethane	0.231	0.258	-11.7	123	0.00
56 T	Ethyl methacrylate	0.258	0.289	-12.0	118	0.00
57 T	1,3-Dichloropropane	0.372	0.397	-6.7	117	0.00
58 T	2-Chloroethyl Vinyl ether	0.068	0.086	-26.5#	127	0.00
59 T	2-Hexanone	0.093	0.108	-16.1	122	0.00
60 T	Dibromochloromethane	0.276	0.303	-9.8	122	0.00
61 T	1,2-Dibromoethane	0.218	0.238	-9.2	121	0.00
62 S	4-Bromofluorobenzene	0.312	0.350	-12.2	125	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	118	0.00
64 T	Tetrachloroethene	0.353	0.346	2.0	119	0.00
65 PM	Chlorobenzene	0.990	1.004	-1.4	117	0.00
66 T	1,1,1,2-Tetrachloroethane	0.336	0.353	-5.1	119	0.00
67 C	Ethyl Benzene	1.755	1.806	-2.9#	115	0.00
68 T	m/p-Xylenes	0.697	0.725	-4.0	116	0.00
69 T	o-Xylene	0.640	0.662	-3.4	114	0.00
70 T	Styrene	1.085	1.151	-6.1	117	0.00
71 P	Bromoform	0.180	0.192	-6.7	120	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
73 T	Isopropylbenzene	3.653	3.607	1.3	115	0.00
74 T	N-amyl acetate	0.593	0.599	-1.0	118	0.00
75 P	1,1,2,2-Tetrachloroethane	0.614	0.615	-0.2	121	0.00
76 T	1,2,3-Trichloropropane	0.388	0.413	-6.4	126	0.00
77 T	Bromobenzene	0.841	0.841	0.0	119	0.00
78 T	n-propylbenzene	4.376	4.337	0.9	114	0.00
79 T	2-Chlorotoluene	2.356	2.330	1.1	116	0.00
80 T	1,3,5-Trimethylbenzene	2.980	2.986	-0.2	115	0.00
81 T	trans-1,4-Dichloro-2-butene	0.163	0.177	-8.6	127	0.00
82 T	4-Chlorotoluene	2.451	2.439	0.5	118	0.00
83 T	tert-Butylbenzene	2.626	2.585	1.6	113	0.00
84 T	1,2,4-Trimethylbenzene	2.941	2.971	-1.0	117	0.00
85 T	sec-Butylbenzene	3.962	3.972	-0.3	117	0.00
86 T	p-Isopropyltoluene	3.337	3.324	0.4	116	0.00
87 T	1,3-Dichlorobenzene	1.744	1.741	0.2	120	0.00
88 T	1,4-Dichlorobenzene	1.715	1.681	2.0	118	0.00
89 T	n-Butylbenzene	3.052	3.071	-0.6	116	0.00
90 T	Hexachloroethane	0.572	0.574	-0.3	116	0.00
91 T	1,2-Dichlorobenzene	1.486	1.476	0.7	119	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.085	0.084	1.2	122	0.00
93 T	1,2,4-Trichlorobenzene	0.917	0.948	-3.4	125	0.00
94 T	Hexachlorobutadiene	0.480	0.504	-5.0	126	0.00
95 T	Naphthalene	1.762	1.788	-1.5	117	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110722\
Data File : VD074751.D
Acq On : 07 Nov 2022 09:53
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Nov 07 11:44:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.807	0.808	-0.1	121	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6