

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
 Data File : VD074888.D
 Acq On : 11 Nov 2022 13:43
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Nov 14 00:54:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	171341	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	258212	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	230755	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	102331	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	56935	50.010	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.020%
35) Dibromofluoromethane	7.810	113	79128	49.713	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.420%
50) Toluene-d8	10.269	98	242493	46.250	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.500%
62) 4-Bromofluorobenzene	12.575	95	73966	45.942	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
Data File : VD074888.D
Acq On : 11 Nov 2022 13:43
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Nov 14 00:54:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

