

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
 Data File : VD074896.D
 Acq On : 11 Nov 2022 17:28
 Operator : VA/SY
 Sample : N5531-15RE
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-27RE

Quant Time: Nov 14 00:56:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	4443	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	8905	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	8635	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	3600	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	4963	183.988	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 367.980%#	
35) Dibromofluoromethane	7.810	113	3852	70.173	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 140.340%	
50) Toluene-d8	10.269	98	10921	60.397	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 120.800%	
62) 4-Bromofluorobenzene	12.575	95	3348	60.298	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 120.600%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
Data File : VD074896.D
Acq On : 11 Nov 2022 17:28
Operator : VA/SY
Sample : N5531-15RE
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-27RE

Quant Time: Nov 14 00:56:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

