

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
 Data File : VD074906.D
 Acq On : 12 Nov 2022 00:26
 Operator : VA/SY
 Sample : N5509-01
 Misc : 4.15G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-7

Quant Time: Nov 14 00:58:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	223858	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	382748	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	328866	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	128550	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	116172	81.879	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	163.760%#
35) Dibromofluoromethane	7.805	113	140011	59.342	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.680%
50) Toluene-d8	10.269	98	435926	56.090	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.180%
62) 4-Bromofluorobenzene	12.575	95	109122	45.725	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
Data File : VD074906.D
Acq On : 12 Nov 2022 00:26
Operator : VA/SY
Sample : N5509-01
Misc : 4.15G/5.00ml/MSVOA_D/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-7

Quant Time: Nov 14 00:58:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

