

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
 Data File : VD074917.D
 Acq On : 12 Nov 2022 05:33
 Operator : VA/SY
 Sample : N5506-01RE
 Misc : 4.69G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-5

Quant Time: Nov 14 01:00:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	118675	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	188435	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	176156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	70141	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	76530	103.376	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	206.760%#
35) Dibromofluoromethane	7.811	113	88473	76.167	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	152.340%#
50) Toluene-d8	10.269	98	249165	65.120	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	130.240%
62) 4-Bromofluorobenzene	12.575	95	66880	56.923	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
Data File : VD074917.D
Acq On : 12 Nov 2022 05:33
Operator : VA/SY
Sample : N5506-01RE
Misc : 4.69G/5.00ml/MSVOA_D/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-5

Quant Time: Nov 14 01:00:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

