

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
 Data File : VD074889.D
 Acq On : 11 Nov 2022 14:11
 Operator : VA/SY
 Sample : VSTDICV050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD111122

Quant Time: Nov 14 00:55:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 11/14/2022
 Supervised By :Mahesh Dadoda 11/14/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	168698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	256491	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	231257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	110530	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	55427	49.373	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	98.740%		
35) Dibromofluoromethane	7.805	113	76407	48.326	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	96.660%		
50) Toluene-d8	10.269	98	249227	47.853	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	95.700%		
62) 4-Bromofluorobenzene	12.575	95	79816	49.908	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	99.820%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.928	85	57501	51.041	ug/l	96
3) Chloromethane	2.146	50	90215	51.105	ug/l	95
4) Vinyl Chloride	2.281	62	117920	49.851	ug/l	100
5) Bromomethane	2.687	94	100458	56.684	ug/l	94
6) Chloroethane	2.834	64	86078	49.554	ug/l	99
7) Trichlorofluoromethane	3.170	101	136258	50.556	ug/l	96
8) Diethyl Ether	3.593	74	40525	52.443	ug/l	98
9) 1,1,2-Trichlorotrifluo...	3.964	101	87286	50.750	ug/l	99
10) Methyl Iodide	4.164	142	100007	48.999	ug/l	99
11) Tert butyl alcohol	5.058	59	21673	267.587	ug/l #	91
12) 1,1-Dichloroethene	3.940	96	87331	52.128	ug/l	93
13) Acrolein	3.793	56	11208	237.132	ug/l	100
14) Allyl chloride	4.558	41	93843	52.368	ug/l	99
15) Acrylonitrile	5.252	53	79463	263.142	ug/l	99
16) Acetone	4.017	43	79943	287.057	ug/l	99
17) Carbon Disulfide	4.264	76	241828	50.741	ug/l	97
18) Methyl Acetate	4.558	43	35750	49.145	ug/l	99
19) Methyl tert-butyl Ether	5.317	73	178882	53.332	ug/l	95
20) Methylene Chloride	4.799	84	126642	54.847	ug/l	92
21) trans-1,2-Dichloroethene	5.311	96	99998	52.636	ug/l	92
22) Diisopropyl ether	6.216	45	217440	54.440	ug/l	96
23) Vinyl Acetate	6.158	43	476602	278.020	ug/l	98
24) 1,1-Dichloroethane	6.111	63	152228	51.036	ug/l	100
25) 2-Butanone	7.081	43	97128	266.706	ug/l	93
26) 2,2-Dichloropropane	7.075	77	142689	50.904	ug/l	99
27) cis-1,2-Dichloroethene	7.081	96	110868	52.603	ug/l	100
28) Bromochloromethane	7.422	49	48052	49.935	ug/l	97
29) Tetrahydrofuran	7.446	42	54826	271.980	ug/l	95
30) Chloroform	7.599	83	169355	51.844	ug/l	98
31) Cyclohexane	7.875	56	119451	48.569	ug/l	99
32) 1,1,1-Trichloroethane	7.793	97	146314	51.972	ug/l	98
36) 1,1-Dichloropropene	8.005	75	112238	48.302	ug/l	99
37) Ethyl Acetate	7.169	43	40975	52.551	ug/l	97
38) Carbon Tetrachloride	7.993	117	105279	48.132	ug/l	98
39) Methylcyclohexane	9.275	83	134395	53.988	ug/l	95
40) Benzene	8.252	78	329934	49.379	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
 Data File : VD074889.D
 Acq On : 11 Nov 2022 14:11
 Operator : VA/SY
 Sample : VSTDICV050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD111122

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 11/14/2022
 Supervised By :Mahesh Dadoda 11/14/2022

Quant Time: Nov 14 00:55:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.405	41	22528m	54.475	ug/l	
42) 1,2-Dichloroethane	8.328	62	77791	48.648	ug/l	99
43) Isopropyl Acetate	8.358	43	69938	50.927	ug/l	96
44) Trichloroethene	9.028	130	92287	52.280	ug/l	97
45) 1,2-Dichloropropane	9.305	63	77383	51.544	ug/l	92
46) Dibromomethane	9.399	93	42863	50.980	ug/l	94
47) Bromodichloromethane	9.587	83	109457	52.456	ug/l	96
48) Methyl methacrylate	9.381	41	32442	53.328	ug/l	93
49) 1,4-Dioxane	9.381	88	8448	1023.964	ug/l	89
51) 4-Methyl-2-Pentanone	10.157	43	183589	280.419	ug/l	96
52) Toluene	10.334	92	216886	54.289	ug/l	100
53) t-1,3-Dichloropropene	10.552	75	99060	53.740	ug/l	100
54) cis-1,3-Dichloropropene	10.016	75	120351	53.239	ug/l	92
55) 1,1,2-Trichloroethane	10.734	97	62096	53.115	ug/l	97
56) Ethyl methacrylate	10.599	69	70019	49.580	ug/l	95
57) 1,3-Dichloropropane	10.881	76	98647	52.485	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.869	63	115856	235.584	ug/l	96
59) 2-Hexanone	10.922	43	132245	289.757	ug/l	98
60) Dibromochloromethane	11.075	129	73040	52.665	ug/l	100
61) 1,2-Dibromoethane	11.181	107	56657	51.819	ug/l	98
64) Tetrachloroethene	10.810	164	75783	48.876	ug/l	95
65) Chlorobenzene	11.610	112	228002	50.399	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.681	131	80715	50.922	ug/l	99
67) Ethyl Benzene	11.687	91	415371	52.701	ug/l	99
68) m/p-Xylenes	11.793	106	328349	104.487	ug/l	99
69) o-Xylene	12.122	106	151066	52.251	ug/l	98
70) Styrene	12.134	104	263166	53.101	ug/l	99
71) Bromoform	12.299	173	40956	49.925	ug/l #	95
73) Isopropylbenzene	12.422	105	402056	52.848	ug/l	99
74) N-amyl acetate	12.234	43	64966	51.903	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.675	83	67454	50.125	ug/l	97
76) 1,2,3-Trichloropropane	12.728	75	45189m	48.881	ug/l	
77) Bromobenzene	12.698	156	88822	49.891	ug/l	94
78) n-propylbenzene	12.763	91	484995	51.863	ug/l	100
79) 2-Chlorotoluene	12.851	91	258690	51.340	ug/l	100
80) 1,3,5-Trimethylbenzene	12.904	105	329653	52.004	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.469	75	18727	49.833	ug/l	95
82) 4-Chlorotoluene	12.946	91	271618	51.764	ug/l	98
83) tert-Butylbenzene	13.169	119	283476	51.538	ug/l	99
84) 1,2,4-Trimethylbenzene	13.210	105	327231	52.195	ug/l	100
85) sec-Butylbenzene	13.346	105	428321	51.195	ug/l	100
86) p-Isopropyltoluene	13.463	119	363313	52.223	ug/l	99
87) 1,3-Dichlorobenzene	13.457	146	187289	50.572	ug/l	99
88) 1,4-Dichlorobenzene	13.540	146	183415	49.900	ug/l	99
89) n-Butylbenzene	13.787	91	330666	51.195	ug/l	99
90) Hexachloroethane	14.051	117	62884	50.469	ug/l	92
91) 1,2-Dichlorobenzene	13.834	146	157624	50.051	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.445	75	8850	47.700	ug/l	91
93) 1,2,4-Trichlorobenzene	15.098	180	97407	50.812	ug/l	99
94) Hexachlorobutadiene	15.204	225	46278	46.801	ug/l	99
95) Naphthalene	15.334	128	184736	51.920	ug/l	99
96) 1,2,3-Trichlorobenzene	15.522	180	82760	49.922	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
Data File : VD074889.D
Acq On : 11 Nov 2022 14:11
Operator : VA/SY
Sample : VSTDICV050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022

Quant Time: Nov 14 00:55:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
 Data File : VD074889.D
 Acq On : 11 Nov 2022 14:11
 Operator : VA/SY
 Sample : VSTDICV050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1


Instrument :
 MSVOA_D
 ClientSampleId :
 ICVD111122

Manual Integrations APPROVED

Reviewed By :Krupa Patel 11/14/2022
 Supervised By :Mahesh Dadoda 11/14/2022


Quant Time: Nov 14 00:55:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVD111122



Tgt Ion:168 Resp: 168698
Ion Ratio Lower Upper
168 100
99 53.4 40.9 61.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#2
Dichlorodifluoromethane
Concen: 51.041 ug/l
RT: 1.928 min Scan# 58
Delta R.T. -0.006 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion: 85 Resp: 57501
Ion Ratio Lower Upper
85 100
87 32.8 17.5 52.6





#3
Chloromethane
Concen: 51.105 ug/l
RT: 2.146 min Scan# 95
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion: 50 Resp: 90215
Ion Ratio Lower Upper
50 100
52 34.2 30.0 45.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#4
Vinyl Chloride
Concen: 49.851 ug/l
RT: 2.281 min Scan# 118
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion: 62 Resp: 117920
Ion Ratio Lower Upper
62 100
64 32.0 25.5 38.3





#5
Bromomethane
Concen: 56.684 ug/l
RT: 2.687 min Scan# 187
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion: 94 Resp: 100458
Ion Ratio Lower Upper
94 100
96 88.0 74.6 112.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#6
Chloroethane
Concen: 49.554 ug/l
RT: 2.834 min Scan# 212
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion: 64 Resp: 86078
Ion Ratio Lower Upper
64 100
66 33.1 25.8 38.8





#7
 Trichlorofluoromethane
 Concen: 50.556 ug/l
 RT: 3.170 min Scan# 269
 Delta R.T. -0.000 min
 Lab File: VD074889.D
 Acq: 11 Nov 2022 14:11




Instrument : MSVOA_D
 ClientSampleId : ICVVD111122



Tgt Ion:101 Resp: 136258
 Ion Ratio Lower Upper
 101 100
 103 65.5 49.8 74.8




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 11/14/2022
 Supervised By :Mahesh Dadoda 11/14/2022



#8
 Diethyl Ether
 Concen: 52.443 ug/l
 RT: 3.593 min Scan# 341
 Delta R.T. 0.000 min
 Lab File: VD074889.D
 Acq: 11 Nov 2022 14:11

Tgt Ion: 74 Resp: 40525
 Ion Ratio Lower Upper
 74 100
 45 77.5 37.9 113.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.750 ug/l
RT: 3.964 min Scan# 404
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122

Tgt Ion:101 Resp: 87286
Ion Ratio Lower Upper
101 100
85 42.5 34.1 51.1
151 76.3 61.7 92.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#10
Methyl Iodide
Concen: 48.999 ug/l
RT: 4.164 min Scan# 438
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Tgt Ion:142 Resp: 100007
Ion Ratio Lower Upper
142 100
127 40.6 31.9 47.9
141 14.4 11.4 17.2









#11
Tert butyl alcohol
Concen: 267.587 ug/l
RT: 5.058 min Scan# 589
Delta R.T. 0.006 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122

Tgt Ion: 59 Resp: 2167
Ion Ratio Lower Upper
59 100
57 8.5 4.4 6.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#12
1,1-Dichloroethene
Concen: 52.128 ug/l
RT: 3.940 min Scan# 400
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion: 96 Resp: 87331
Ion Ratio Lower Upper
96 100
61 132.8 113.9 170.9
98 61.4 51.8 77.8





#13
Acrolein
Concen: 237.132 ug/l
RT: 3.793 min Scan# 31
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion: 56 Resp: 11208
Ion Ratio Lower Upper
56 100
55 67.7 54.0 81.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#14
Allyl chloride
Concen: 52.368 ug/l
RT: 4.558 min Scan# 505
Delta R.T. -0.006 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion: 41 Resp: 93843
Ion Ratio Lower Upper
41 100
39 67.6 53.4 80.2
76 53.6 42.4 63.6





#15
Acrylonitrile
Concen: 263.142 ug/l
RT: 5.252 min Scan# 61
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122




Tgt Ion: 53 Resp: 7946
Ion Ratio Lower Upper
53 100
52 82.0 66.0 99.0
51 35.7 28.3 42.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022




#16
Acetone
Concen: 287.057 ug/l
RT: 4.017 min Scan# 413
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Tgt Ion: 43 Resp: 79943
Ion Ratio Lower Upper
43 100
58 37.1 30.2 45.2





#17
Carbon Disulfide
Concen: 50.741 ug/l
RT: 4.264 min Scan# 41
Delta R.T. -0.006 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion: 76 Resp: 241828
Ion Ratio Lower Upper
76 100
78 8.9 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#18
Methyl Acetate
Concen: 49.145 ug/l
RT: 4.558 min Scan# 505
Delta R.T. -0.006 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Tgt Ion: 43 Resp: 35750
Ion Ratio Lower Upper
43 100
74 31.7 24.7 37.1





#19

Methyl tert-butyl Ether

Concen: 53.332 ug/l

RT: 5.317 min Scan# 61

Delta R.T. -0.000 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11

Instrument :

MSVOA_D

ClientSampleId :

ICVVD111122

Tgt Ion: 73 Resp: 17888

Ion Ratio Lower Upper

73 100





57 22.3 16.1 24.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/14/2022

Supervised By :Mahesh Dadoda 11/14/2022



#20

Methylene Chloride

Concen: 54.847 ug/l

RT: 4.799 min Scan# 546

Delta R.T. -0.000 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11

Tgt Ion: 84 Resp: 126642




Ion Ratio Lower Upper


84 100

49 94.6 68.0 102.0

51 31.5 23.2 34.8


86 65.5 48.0 72.0





#21
trans-1,2-Dichloroethene
Concen: 52.636 ug/l
RT: 5.311 min Scan# 611
Delta R.T. 0.006 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion: 96 Resp: 99998
Ion Ratio Lower Upper
96 100
61 111.0 95.7 143.5
98 62.0 54.8 82.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#22
Diisopropyl ether
Concen: 54.440 ug/l
RT: 6.216 min Scan# 787
Delta R.T. 0.005 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Tgt Ion: 45 Resp: 217440
Ion Ratio Lower Upper
45 100
43 50.6 37.8 56.6
87 34.8 26.6 40.0
59 13.0 9.9 14.9









#23
Vinyl Acetate
Concen: 278.020 ug/l
RT: 6.158 min Scan# 71
Delta R.T. 0.006 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122

Tgt Ion: 43 Resp: 476602
Ion Ratio Lower Upper
43 100
86 13.1 11.2 16.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#24
1,1-Dichloroethane
Concen: 51.036 ug/l
RT: 6.111 min Scan# 769
Delta R.T. -0.006 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion: 63 Resp: 152228
Ion Ratio Lower Upper
63 100
98 9.2 4.6 13.8
100 5.7 2.9 8.6





#25
2-Butanone
Concen: 266.706 ug/l
RT: 7.081 min Scan# 912
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion: 43 Resp: 97128
Ion Ratio Lower Upper
43 100
72 36.1 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#26
2,2-Dichloropropane
Concen: 50.904 ug/l
RT: 7.075 min Scan# 933
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion: 77 Resp: 142689
Ion Ratio Lower Upper
77 100
97 24.7 12.5 37.5





#27
 cis-1,2-Dichloroethene
 Concen: 52.603 ug/l
 RT: 7.081 min Scan# 911
 Delta R.T. 0.000 min
 Lab File: VD074889.D
 Acq: 11 Nov 2022 14:11




Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD111122



Tgt Ion: 96 Resp: 11086
 Ion Ratio Lower Upper
 96 100
 61 121.6 0.0 242.4
 98 64.5 0.0 130.0




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 11/14/2022
 Supervised By :Mahesh Dadoda 11/14/2022



#28
 Bromochloromethane
 Concen: 49.935 ug/l
 RT: 7.422 min Scan# 992
 Delta R.T. -0.006 min
 Lab File: VD074889.D
 Acq: 11 Nov 2022 14:11

Tgt Ion: 49 Resp: 48052
 Ion Ratio Lower Upper
 49 100
 129 2.9 0.0 5.0
 130 101.6 84.1 126.1





#29

Tetrahydrofuran

Concen: 271.980 ug/l

RT: 7.446 min Scan# 996

Delta R.T. -0.006 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11

Instrument :

MSVOA_D

ClientSampleId :

ICVVD111122





Tgt Ion:	42	Resp:	54820
Ion	Ratio	Lower	Upper
42	100		
72	54.7	45.8	68.6
71	51.6	44.6	66.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/14/2022

Supervised By :Mahesh Dadoda 11/14/2022



#30

Chloroform

Concen: 51.844 ug/l




RT: 7.599 min Scan# 1022


Delta R.T. 0.006 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11


Tgt Ion:	83	Resp:	169355
Ion	Ratio	Lower	Upper
83	100		
85	63.4	52.0	78.0





#31
Cyclohexane
Concen: 48.569 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion: 56 Resp: 11945
Ion Ratio Lower Upper
56 100
69 36.8 29.2 43.8
84 104.9 83.4 125.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#32
1,1,1-Trichloroethane
Concen: 51.972 ug/l
RT: 7.793 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Tgt Ion: 97 Resp: 146314
Ion Ratio Lower Upper
97 100
99 64.3 53.3 79.9
61 38.0 30.2 45.4





#33

1,2-Dichloroethane-d4

Concen: 49.373 ug/l

RT: 8.228 min Scan# 1129

Delta R.T. 0.000 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11

Instrument :

MSVOA_D

ClientSampleId :

ICVVD111122

Tgt Ion: 65 Resp: 5542

Ion Ratio Lower Upper

65 100





67 53.6 0.0 110.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/14/2022

Supervised By :Mahesh Dadoda 11/14/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1222

Delta R.T. 0.000 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11




Tgt Ion:114 Resp: 256491


Ion Ratio Lower Upper

114 100

63 16.2 0.0 33.2

88 13.6 0.0 28.2





#35

Dibromofluoromethane

Concen: 48.326 ug/l

RT: 7.805 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11

Instrument :

MSVOA_D

ClientSampleId :

ICVVD111122





Tgt Ion:	113	Resp:	7640
Ion	Ratio	Lower	Upper
113	100		
111	103.9	82.3	123.5
192	19.6	16.4	24.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/14/2022

Supervised By :Mahesh Dadoda 11/14/2022



#36

1,1-Dichloropropene

Concen: 48.302 ug/l




RT: 8.005 min Scan# 1091


Delta R.T. -0.006 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11

Tgt Ion:	75	Resp:	112238
Ion	Ratio	Lower	Upper
75	100		
110	38.2	19.7	59.0
77	31.2	25.4	38.2





#37

Ethyl Acetate

Concen: 52.551 ug/l

RT: 7.169 min Scan# 948

Delta R.T. 0.005 min

Lab File: VD074889.D


Acq: 11 Nov 2022 14:11

Instrument :

MSVOA_D

ClientSampleId :

ICVVD111122






Tgt Ion:	43	Resp:	40975
Ion Ratio	Lower	Upper	
43	100		
61	17.0	12.2	18.4
70	12.4	9.6	14.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/14/2022

Supervised By :Mahesh Dadoda 11/14/2022



#38

Carbon Tetrachloride

Concen: 48.132 ug/l




RT: 7.993 min Scan# 1089


Delta R.T. 0.006 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11

Tgt Ion:	117	Resp:	105279
Ion Ratio	Lower	Upper	
117	100		
119	93.5	73.1	109.7
121	30.0	25.0	37.6





#39

Methylcyclohexane

Concen: 53.988 ug/l

RT: 9.275 min Scan# 1307

Delta R.T. 0.000 min

Lab File: VD074889.D


Acq: 11 Nov 2022 14:11

Instrument :

MSVOA_D

ClientSampleId :

ICVVD111122






Tgt Ion:	83	Resp:	13439
Ion Ratio	Lower	Upper	
83	100		
55	65.6	49.0	73.4
98	49.6	41.7	62.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/14/2022

Supervised By :Mahesh Dadoda 11/14/2022



#40

Benzene

Concen: 49.379 ug/l




RT: 8.252 min Scan# 1133


Delta R.T. -0.000 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11


Tgt Ion:	78	Resp:	329934
Ion Ratio	Lower	Upper	
78	100		
77	24.8	20.1	30.1





#41
Methacrylonitrile
Concen: 54.475 ug/l m
RT: 7.405 min Scan# 989
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion: 41 Resp: 22528
Ion Ratio Lower Upper
41 100
39 53.3 35.0 52.6
67 94.7 77.0 115.6
52 32.1 29.7 44.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#42
1,2-Dichloroethane
Concen: 48.648 ug/l
RT: 8.328 min Scan# 1146
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Tgt Ion: 62 Resp: 77791
Ion Ratio Lower Upper
62 100
98 11.7 0.0 24.2





#43

Isopropyl Acetate

Concen: 50.927 ug/l

RT: 8.358 min Scan# 1151

Delta R.T. -0.005 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11

Instrument :

MSVOA_D

ClientSampleId :

ICVVD111122

Tgt Ion: 43 Resp: 69938





Ion	Ratio	Lower	Upper
43	100		
61	35.6	30.9	46.3
87	16.7	14.3	21.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/14/2022

Supervised By :Mahesh Dadoda 11/14/2022



#44

Trichloroethene

Concen: 52.280 ug/l

RT: 9.028 min Scan# 1265




Delta R.T. 0.000 min


Lab File: VD074889.D

Acq: 11 Nov 2022 14:11

Tgt Ion:130 Resp: 92287

Ion	Ratio	Lower	Upper
130	100		
95	96.0	0.0	185.4





#45

1,2-Dichloropropane

Concen: 51.544 ug/l

RT: 9.305 min Scan# 1312

Delta R.T. -0.000 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11

Instrument :

MSVOA_D

ClientSampleId :

ICVVD111122

Tgt Ion: 63 Resp: 7738

Ion Ratio Lower Upper

63 100





65 33.1 23.1 34.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/14/2022

Supervised By :Mahesh Dadoda 11/14/2022



#46

Dibromomethane

Concen: 50.980 ug/l

RT: 9.399 min Scan# 1328

Delta R.T. 0.006 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11




Tgt Ion: 93 Resp: 42863


Ion Ratio Lower Upper

93 100

95 84.3 68.2 102.2


174 98.8 87.4 131.0





#47
Bromodichloromethane
Concen: 52.456 ug/l
RT: 9.587 min Scan# 1360
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion: 83 Resp: 10945
Ion Ratio Lower Upper
83 100
85 64.6 48.8 73.2
127 8.6 6.7 10.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#48
Methyl methacrylate
Concen: 53.328 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Tgt Ion: 41 Resp: 32442
Ion Ratio Lower Upper
41 100
69 109.0 93.0 139.4
39 51.0 44.6 66.8





#49

1,4-Dioxane

Concen: 1023.964 ug/l

RT: 9.381 min Scan# 1325

Delta R.T. 0.000 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11

Instrument :

MSVOA_D

ClientSampleId :

ICVVD111122

Tgt Ion: 88 Resp: 844

Ion Ratio Lower Upper

88 100

43 28.1 24.1 36.1





58 64.4 43.1 64.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/14/2022

Supervised By :Mahesh Dadoda 11/14/2022



#50

Toluene-d8

Concen: 47.853 ug/l

RT: 10.269 min Scan# 1476

Delta R.T. 0.000 min

Lab File: VD074889.D




Acq: 11 Nov 2022 14:11


Tgt Ion: 98 Resp: 249227

Ion Ratio Lower Upper

98 100


100 64.5 51.4 77.0





#51
4-Methyl-2-Pentanone
Concen: 280.419 ug/l
RT: 10.157 min Scan# 1457
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122




Tgt Ion: 43 Resp: 183589
Ion Ratio Lower Upper
43 100
58 44.8 37.8 56.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022




#52
Toluene
Concen: 54.289 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Tgt Ion: 92 Resp: 216886
Ion Ratio Lower Upper
92 100
91 174.0 139.0 208.4





#53
t-1,3-Dichloropropene
Concen: 53.740 ug/l
RT: 10.552 min Scan# 1525
Delta R.T. -0.005 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument : MSVOA_D
ClientSampleId : ICVVD111122



Tgt Ion: 75 Resp: 99060
Ion Ratio Lower Upper
75 100
77 31.8 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#54
cis-1,3-Dichloropropene
Concen: 53.239 ug/l
RT: 10.016 min Scan# 1433
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion: 75 Resp: 120351
Ion Ratio Lower Upper
75 100
77 35.7 24.7 37.1
39 36.4 25.4 38.2





#55
1,1,2-Trichloroethane
Concen: 53.115 ug/l
RT: 10.734 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion:	97	Resp:	62090
Ion Ratio	100	Lower	Upper
97	100		
83	87.1	64.8	97.2
85	54.6	44.8	67.2
99	61.9	49.4	74.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#56
Ethyl methacrylate
Concen: 49.580 ug/l
RT: 10.599 min Scan# 1532
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Tgt Ion:	69	Resp:	70019
Ion Ratio	100	Lower	Upper
69	100		
41	52.7	38.6	58.0
39	24.1	19.9	29.9









#57
1,3-Dichloropropane
Concen: 52.485 ug/l
RT: 10.881 min Scan# 1580
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Instrument : MSVOA_D
ClientSampleId : ICVVD111122

Tgt Ion: 76 Resp: 9864
Ion Ratio Lower Upper
76 100
78 32.0 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#58
2-Chloroethyl Vinyl ether
Concen: 235.584 ug/l
RT: 9.869 min Scan# 1408
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion: 63 Resp: 115856
Ion Ratio Lower Upper
63 100
106 31.0 26.5 39.7





#59
2-Hexanone
Concen: 289.757 ug/l
RT: 10.922 min Scan# 1587
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion: 43 Resp: 132245
Ion Ratio Lower Upper
43 100
58 62.7 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#60
Dibromochloromethane
Concen: 52.665 ug/l
RT: 11.075 min Scan# 1613
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Tgt Ion:129 Resp: 73040
Ion Ratio Lower Upper
129 100
127 78.0 39.1 117.2









#61
1,2-Dibromoethane
Concen: 51.819 ug/l
RT: 11.181 min Scan# 1631
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122

Tgt Ion:107 Resp: 5665
Ion Ratio Lower Upper
107 100
109 95.1 74.6 111.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#62
4-Bromofluorobenzene
Concen: 49.908 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion: 95 Resp: 79816
Ion Ratio Lower Upper
95 100
174 90.1 0.0 197.0
176 86.7 0.0 182.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1699
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion:117 Resp: 23125
Ion Ratio Lower Upper
117 100
82 51.3 41.6 62.4
119 31.8 25.2 37.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#64
Tetrachloroethene
Concen: 48.876 ug/l
RT: 10.810 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion:164 Resp: 75783
Ion Ratio Lower Upper
164 100
166 128.0 107.8 161.8
129 90.3 77.4 116.2
131 88.2 69.5 104.3





#65
Chlorobenzene
Concen: 50.399 ug/l
RT: 11.610 min Scan# 1704
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion:112 Resp: 22800
Ion Ratio Lower Upper
112 100
114 32.2 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 50.922 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Tgt Ion:131 Resp: 80715
Ion Ratio Lower Upper
131 100
133 93.5 46.6 139.8
119 67.4 32.5 97.4









#67
Ethyl Benzene
Concen: 52.701 ug/l
RT: 11.687 min Scan# 1717
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122

Tgt Ion: 91 Resp: 41537
Ion Ratio Lower Upper
91 100
106 32.6 26.6 40.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#68
m/p-Xylenes
Concen: 104.487 ug/l
RT: 11.793 min Scan# 1735
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Tgt Ion:106 Resp: 328349
Ion Ratio Lower Upper
106 100
91 193.2 156.2 234.4









#69
o-Xylene
Concen: 52.251 ug/l
RT: 12.122 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122

Tgt Ion:106 Resp: 151060
Ion Ratio Lower Upper
106 100
91 205.7 104.5 313.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#70
Styrene
Concen: 53.101 ug/l
RT: 12.134 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion:104 Resp: 263166
Ion Ratio Lower Upper
104 100
78 43.8 35.7 53.5
103 55.7 43.7 65.5





#71
Bromoform
Concen: 49.925 ug/l
RT: 12.299 min Scan# 1821
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122




Tgt Ion:173 Resp: 40950
Ion Ratio Lower Upper
173 100
175 49.7 23.1 69.3
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Tgt Ion:152 Resp: 110530
Ion Ratio Lower Upper
152 100
115 55.6 26.8 80.4
150 173.6 0.0 348.2





#73
Isopropylbenzene
Concen: 52.848 ug/l
RT: 12.422 min Scan# 1842
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion:105 Resp: 402050
Ion Ratio Lower Upper
105 100
120 26.4 13.5 40.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#74
N-ethyl acetate
Concen: 51.903 ug/l
RT: 12.234 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion: 43 Resp: 64966
Ion Ratio Lower Upper
43 100
70 55.7 43.8 65.6
55 29.0 22.0 33.0
61 28.9 23.9 35.9





#75
1,1,2,2-Tetrachloroethane
Concen: 50.125 ug/l
RT: 12.675 min Scan# 1885
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument : MSVOA_D
ClientSampleId : ICVVD111122




Tgt Ion: 83 Resp: 67454
Ion Ratio Lower Upper
83 100
131 10.2 4.7 14.1
85 63.2 33.1 99.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022




#76
1,2,3-Trichloropropane
Concen: 48.881 ug/l m
RT: 12.728 min Scan# 1894
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Tgt Ion: 75 Resp: 45189
Ion Ratio Lower Upper
75 100
77 0.0 185.8 557.5#





#77
Bromobenzene
Concen: 49.891 ug/l
RT: 12.698 min Scan# 1889
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122




Tgt Ion:156 Resp: 8882
Ion Ratio Lower Upper
156 100
77 168.7 77.7 233.1
158 99.2 49.4 148.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022




#78
n-propylbenzene
Concen: 51.863 ug/l
RT: 12.763 min Scan# 1900
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11



Tgt Ion: 91 Resp: 484995
Ion Ratio Lower Upper
91 100
120 23.6 11.9 35.7









#79
2-Chlorotoluene
Concen: 51.340 ug/l
RT: 12.851 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11

Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122

Tgt Ion: 91 Resp: 258690
Ion Ratio Lower Upper
91 100
126 36.2 18.3 54.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#80
1,3,5-Trimethylbenzene
Concen: 52.004 ug/l
RT: 12.904 min Scan# 1924
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion:105 Resp: 329653
Ion Ratio Lower Upper
105 100
120 50.3 25.7 77.0





#81
trans-1,4-Dichloro-2-butene
Concen: 49.833 ug/l
RT: 12.469 min Scan# 1850
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion: 75 Resp: 1872
Ion Ratio Lower Upper
75 100
53 75.2 63.8 95.6
89 44.3 38.2 57.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#82
4-Chlorotoluene
Concen: 51.764 ug/l
RT: 12.946 min Scan# 1931
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion: 91 Resp: 271618
Ion Ratio Lower Upper
91 100
126 36.7 18.8 56.4





#83
tert-Butylbenzene
Concen: 51.538 ug/l
RT: 13.169 min Scan# 1969
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122




Tgt Ion:119 Resp: 283470
Ion Ratio Lower Upper
119 100
91 59.3 30.0 90.0
134 26.9 13.4 40.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022




#84
1,2,4-Trimethylbenzene
Concen: 52.195 ug/l
RT: 13.210 min Scan# 1976
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Tgt Ion:105 Resp: 327231
Ion Ratio Lower Upper
105 100
120 46.1 22.9 68.7





#85
 sec-Butylbenzene
 Concen: 51.195 ug/l
 RT: 13.346 min Scan# 1999
 Delta R.T. -0.000 min
 Lab File: VD074889.D
 Acq: 11 Nov 2022 14:11




Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD111122




Tgt Ion:105 Resp: 42832
 Ion Ratio Lower Upper
 105 100
 134 20.2 10.0 30.0

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 11/14/2022
 Supervised By :Mahesh Dadoda 11/14/2022




#86
 p-Isopropyltoluene
 Concen: 52.223 ug/l
 RT: 13.463 min Scan# 2019
 Delta R.T. 0.000 min
 Lab File: VD074889.D
 Acq: 11 Nov 2022 14:11




Tgt Ion:119 Resp: 363313
 Ion Ratio Lower Upper
 119 100
 134 25.9 13.5 40.4
 91 21.5 10.8 32.4





#87
1,3-Dichlorobenzene
Concen: 50.572 ug/l
RT: 13.457 min Scan# 2018
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122




Tgt Ion:146 Resp: 187289
Ion Ratio Lower Upper
146 100
111 37.8 18.3 54.8
148 63.3 31.8 95.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022




#88
1,4-Dichlorobenzene
Concen: 49.900 ug/l
RT: 13.540 min Scan# 2032
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11



Tgt Ion:146 Resp: 183415
Ion Ratio Lower Upper
146 100
111 36.3 17.7 53.1
148 63.5 32.3 96.9





#89

n-Butylbenzene

Concen: 51.195 ug/l

RT: 13.787 min Scan# 2074

Delta R.T. -0.000 min

Lab File: VD074889.D

Acq: 11 Nov 2022 14:11

Instrument :

MSVOA_D

ClientSampleId :

ICVVD111122

Tgt Ion: 91 Resp: 33066





Ion	Ratio	Lower	Upper
91	100		
92	53.1	26.9	80.7
134	26.0	13.4	40.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/14/2022

Supervised By :Mahesh Dadoda 11/14/2022



#90

Hexachloroethane

Concen: 50.469 ug/l

RT: 14.051 min Scan# 2119




Delta R.T. 0.000 min


Lab File: VD074889.D

Acq: 11 Nov 2022 14:11

Tgt Ion:117 Resp: 62884


Ion	Ratio	Lower	Upper
117	100		
201	69.5	38.0	114.0





#91
1,2-Dichlorobenzene
Concen: 50.051 ug/l
RT: 13.834 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122



Tgt Ion:146 Resp: 157624
Ion Ratio Lower Upper
146 100
111 39.6 18.9 56.5
148 63.6 31.9 95.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.700 ug/l
RT: 14.445 min Scan# 2186
Delta R.T. -0.006 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11


Tgt Ion: 75 Resp: 8850
Ion Ratio Lower Upper
75 100
155 97.7 51.2 153.8
157 121.0 68.2 204.5





#93
1,2,4-Trichlorobenzene
Concen: 50.812 ug/l
RT: 15.098 min Scan# 2297
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122




Tgt Ion:180 Resp: 9740
Ion Ratio Lower Upper
180 100
182 92.9 46.9 140.6
145 28.8 14.1 42.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022




#94
Hexachlorobutadiene
Concen: 46.801 ug/l
RT: 15.204 min Scan# 2315
Delta R.T. 0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Tgt Ion:225 Resp: 46278
Ion Ratio Lower Upper
225 100
223 64.1 31.9 95.5
227 64.1 32.6 97.8





#95
Naphthalene
Concen: 51.920 ug/l
RT: 15.334 min Scan# 2337
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11




Instrument :
MSVOA_D
ClientSampleId :
ICVVD111122




Tgt Ion:128 Resp: 184730
Ion Ratio Lower Upper
128 100
127 12.8 10.6 16.0
129 10.8 8.3 12.5

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 11/14/2022
Supervised By :Mahesh Dadoda 11/14/2022



#96
1,2,3-Trichlorobenzene
Concen: 49.922 ug/l
RT: 15.522 min Scan# 2369
Delta R.T. -0.000 min
Lab File: VD074889.D
Acq: 11 Nov 2022 14:11



Tgt Ion:180 Resp: 82760
Ion Ratio Lower Upper
180 100
182 95.2 46.7 140.1
145 31.1 14.9 44.7

