

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
 Data File : VD074910.D
 Acq On : 12 Nov 2022 02:18
 Operator : VA/SY
 Sample : N5607-01
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-01

Quant Time: Nov 14 00:59:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	14623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	24796	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	24943	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	9601	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	16114	181.414	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	362.820%#
35) Dibromofluoromethane	7.804	113	16221	106.124	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	212.240%#
50) Toluene-d8	10.269	98	34015	67.558	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	135.120%
62) 4-Bromofluorobenzene	12.575	95	10067	65.114	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	130.220%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
Data File : VD074910.D
Acq On : 12 Nov 2022 02:18
Operator : VA/SY
Sample : N5607-01
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-01

Quant Time: Nov 14 00:59:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. -0.000 min
Lab File: VD074910.D
Acq: 12 Nov 2022 02:18




Instrument :
MSVOA_D
ClientSampleId :
SB-01

Tgt Ion:168 Resp: 14623
Ion Ratio Lower Upper
168 100
99 51.8 40.9 61.3



#33
1,2-Dichloroethane-d4
Concen: 181.414 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VD074910.D
Acq: 12 Nov 2022 02:18

Tgt Ion: 65 Resp: 16114
Ion Ratio Lower Upper
65 100
67 57.1 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD074910.D

Acq: 12 Nov 2022 02:18

Instrument :

MSVOA_D

ClientSampleId :

SB-01




Tgt Ion:114 Resp: 24796

Ion Ratio Lower Upper

114 100

63 15.4 0.0 33.2

88 14.9 0.0 28.2



#35

Dibromofluoromethane

Concen: 106.124 ug/l

RT: 7.804 min Scan# 1057

Delta R.T. -0.001 min

Lab File: VD074910.D

Acq: 12 Nov 2022 02:18




Tgt Ion:113 Resp: 16221


Ion Ratio Lower Upper

113 100

111 103.7 82.3 123.5

192 19.0 16.4 24.6








#50
Toluene-d8
Concen: 67.558 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VD074910.D
Acq: 12 Nov 2022 02:18




Instrument :
MSVOA_D
ClientSampleId :
SB-01


Tgt Ion: 98 Resp: 34015
Ion Ratio Lower Upper
98 100
100 65.5 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 65.114 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD074910.D
Acq: 12 Nov 2022 02:18

Tgt Ion: 95 Resp: 10067
Ion Ratio Lower Upper
95 100
174 103.4 0.0 197.0
176 95.1 0.0 182.6








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1
Delta R.T. -0.000 min
Lab File: VD074910.D
Acq: 12 Nov 2022 02:18

Instrument :
MSVOA_D
ClientSampleId :
SB-01

Tgt Ion:117 Resp: 24943
Ion Ratio Lower Upper
117 100
82 51.5 41.6 62.4
119 29.6 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.000 min
Lab File: VD074910.D
Acq: 12 Nov 2022 02:18

Tgt Ion:152 Resp: 9601
Ion Ratio Lower Upper
152 100
115 55.0 26.8 80.4
150 164.9 0.0 348.2

