



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111122\
 Data File : VD074901.D
 Acq On : 11 Nov 2022 21:11
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Title : SW846 8260
 Last Update : Mon Nov 14 00:51:58 2022



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.7	8316	PASS
75	95	30	60	47.7	25312	PASS
95	95	100	100	100.0	53077	PASS
96	95	5	9	6.6	3504	PASS
173	174	0.00	2	1.7	878	PASS
174	95	50	100	97.2	51579	PASS
175	174	5	9	7.3	3753	PASS
176	174	95	101	95.6	49291	PASS
177	176	5	9	6.3	3113	PASS