

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111722\
 Data File : VD074965.D
 Acq On : 17 Nov 2022 13:12
 Operator : VA/SY
 Sample : N5664-02
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-2

Quant Time: Nov 17 22:35:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	103656	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	198587	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	168751	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	54096	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	66803	103.308	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	206.620%#
35) Dibromofluoromethane	7.811	113	77589	63.382	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	126.760%
50) Toluene-d8	10.275	98	232372	57.626	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.260%
62) 4-Bromofluorobenzene	12.575	95	60518	48.875	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111722\
Data File : VD074965.D
Acq On : 17 Nov 2022 13:12
Operator : VA/SY
Sample : N5664-02
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-2

Quant Time: Nov 17 22:35:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

