

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111722\
 Data File : VD074967.D
 Acq On : 17 Nov 2022 14:07
 Operator : VA/SY
 Sample : N5642-01RE
 Misc : 5.96G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SOIL-PILE-21RE

Quant Time: Nov 17 22:35:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	129002	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	234667	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	215800	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	85787	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	69210	84.875	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	169.740%#
35) Dibromofluoromethane	7.810	113	82800	57.239	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.480%
50) Toluene-d8	10.269	98	292400	61.364	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	122.720%
62) 4-Bromofluorobenzene	12.575	95	80227	54.831	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111722\
Data File : VD074967.D
Acq On : 17 Nov 2022 14:07
Operator : VA/SY
Sample : N5642-01RE
Misc : 5.96G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SOIL-PILE-21RE

Quant Time: Nov 17 22:35:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

