

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111722\
 Data File : VD074969.D
 Acq On : 17 Nov 2022 15:01
 Operator : VA/SY
 Sample : N5642-08RE
 Misc : 6.53G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RIGHT-SIDE-FLATS-28RE

Quant Time: Nov 17 22:36:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	2261	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	3913	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	3771	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	1653	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	1567	111.603	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 223.200%#	
35) Dibromofluoromethane	7.805	113	1507	62.477	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 124.960%	
50) Toluene-d8	10.269	98	4743	59.694	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 119.380%	
62) 4-Bromofluorobenzene	12.575	95	1510	61.890	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 123.780%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111722\
Data File : VD074969.D
Acq On : 17 Nov 2022 15:01
Operator : VA/SY
Sample : N5642-08RE
Misc : 6.53G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RIGHT-SIDE-FLATS-28RE

Quant Time: Nov 17 22:36:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

