

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111722\
 Data File : VD074971.D
 Acq On : 17 Nov 2022 15:56
 Operator : VA/SY
 Sample : N5642-11
 Misc : 6.69G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 LEFT-SIDE-FLATS-31

Quant Time: Nov 17 22:36:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	128595	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	235005	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	219458	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	95513	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	61919	75.485	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	150.980%
35) Dibromofluoromethane	7.810	113	80751	55.743	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.480%
50) Toluene-d8	10.269	98	293187	61.440	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	122.880%
62) 4-Bromofluorobenzene	12.575	95	83388	56.909	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111722\
Data File : VD074971.D
Acq On : 17 Nov 2022 15:56
Operator : VA/SY
Sample : N5642-11
Misc : 6.69G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
LEFT-SIDE-FLATS-31

Quant Time: Nov 17 22:36:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

