

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111722\
 Data File : VD074973.D
 Acq On : 17 Nov 2022 16:50
 Operator : VA/SY
 Sample : N5642-13RE
 Misc : 6.58G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 LEFT-SIDE-FLATS-33RE

Quant Time: Nov 17 22:36:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	127838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	232184	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	219327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	97292	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	66333	81.867	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 163.740%#	
35) Dibromofluoromethane	7.810	113	82519	57.655	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 115.320%	
50) Toluene-d8	10.275	98	303918	64.463	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 128.920%	
62) 4-Bromofluorobenzene	12.575	95	86698	59.887	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 119.780%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.805	84	34273	7.993	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111722\
Data File : VD074973.D
Acq On : 17 Nov 2022 16:50
Operator : VA/SY
Sample : N5642-13RE
Misc : 6.58G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
LEFT-SIDE-FLATS-33RE

Quant Time: Nov 17 22:36:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

