

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111722\
 Data File : VD074979.D
 Acq On : 17 Nov 2022 19:33
 Operator : VA/SY
 Sample : N5678-07
 Misc : 7.46G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 250FB-BP-BACKFILL-04

Quant Time: Nov 17 22:37:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	129986	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	244899	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	227033	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	96710	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	69388	84.415	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	168.840%#
35) Dibromofluoromethane	7.810	113	88289	58.484	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.960%
50) Toluene-d8	10.269	98	316549	63.656	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	127.320%
62) 4-Bromofluorobenzene	12.575	95	88581	58.011	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.020%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111722\
Data File : VD074979.D
Acq On : 17 Nov 2022 19:33
Operator : VA/SY
Sample : N5678-07
Misc : 7.46G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
250FB-BP-BACKFILL-04

Quant Time: Nov 17 22:37:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.881 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VD074979.D
Acq: 17 Nov 2022 19:33




Instrument :
MSVOA_D
ClientSampleId :
250FB-BP-BACKFILL-04


Tgt Ion:168 Resp: 129986
Ion Ratio Lower Upper
168 100
99 51.2 40.9 61.3



#33
1,2-Dichloroethane-d4
Concen: 84.415 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VD074979.D
Acq: 17 Nov 2022 19:33

Tgt Ion: 65 Resp: 69388
Ion Ratio Lower Upper
65 100
67 56.6 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD074979.D

Acq: 17 Nov 2022 19:33

Instrument :

MSVOA_D

ClientSampleId :

250FB-BP-BACKFILL-04





Tgt Ion:114 Resp: 244899

Ion Ratio Lower Upper

114 100

63 16.0 0.0 33.2

88 13.4 0.0 28.2



#35

Dibromofluoromethane

Concen: 58.484 ug/l

RT: 7.810 min Scan# 1058

Delta R.T. 0.005 min

Lab File: VD074979.D

Acq: 17 Nov 2022 19:33




Tgt Ion:113 Resp: 88289


Ion Ratio Lower Upper

113 100

111 103.0 82.3 123.5

192 19.1 16.4 24.6









#50
Toluene-d8
Concen: 63.656 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD074979.D
Acq: 17 Nov 2022 19:33




Instrument :
MSVOA_D
ClientSampleId :
250FB-BP-BACKFILL-04


Tgt Ion: 98 Resp: 316549
Ion Ratio Lower Upper
98 100
100 64.4 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 58.011 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD074979.D
Acq: 17 Nov 2022 19:33

Tgt Ion: 95 Resp: 88581
Ion Ratio Lower Upper
95 100
174 95.7 0.0 197.0
176 90.9 0.0 182.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.587 min Scan# 11
Delta R.T. 0.006 min
Lab File: VD074979.D
Acq: 17 Nov 2022 19:33

Instrument :
MSVOA_D
ClientSampleId :
250FB-BP-BACKFILL-04

Tgt Ion:117 Resp: 227033
Ion Ratio Lower Upper
117 100
82 51.1 41.6 62.4
119 33.3 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.522 min Scan# 2029
Delta R.T. 0.006 min
Lab File: VD074979.D
Acq: 17 Nov 2022 19:33

Tgt Ion:152 Resp: 96710
Ion Ratio Lower Upper
152 100
115 55.4 26.8 80.4
150 156.6 0.0 348.2

