

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD112122\
 Data File : VD075006.D
 Acq On : 21 Nov 2022 16:20
 Operator : VA/SY
 Sample : N5740-01
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-12

Quant Time: Nov 22 00:18:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	11025	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	20822	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	23427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	12162	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	11177	166.360	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	332.720%#
35) Dibromofluoromethane	7.810	113	9789	76.266	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	152.540%#
50) Toluene-d8	10.269	98	23775	56.232	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.460%
62) 4-Bromofluorobenzene	12.575	95	9370	72.172	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	144.340%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD112122\
Data File : VD075006.D
Acq On : 21 Nov 2022 16:20
Operator : VA/SY
Sample : N5740-01
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-12

Quant Time: Nov 22 00:18:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

