

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111222\
 Data File : VD075009.D
 Acq On : 21 Nov 2022 17:44
 Operator : VA/SY
 Sample : N5695-01
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HD-1-111722

Quant Time: Nov 22 00:19:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	38358	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	73286	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	74806	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	32367	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	32600	138.378	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 276.760%#		
35) Dibromofluoromethane	7.810	113	31500	69.728	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 139.460%		
50) Toluene-d8	10.269	98	85360	57.361	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 114.720%		
62) 4-Bromofluorobenzene	12.575	95	27048	59.193	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 118.380%		
Target Compounds						
20) Methylene Chloride	4.811	84	10868	9.472	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD112122\
Data File : VD075009.D
Acq On : 21 Nov 2022 17:44
Operator : VA/SY
Sample : N5695-01
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HD-1-111722

Quant Time: Nov 22 00:19:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

