

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122822\
 Data File : VD075060.D
 Acq On : 28 Dec 2022 12:34
 Operator : KP/SY
 Sample : VD1228SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1228SBL01

Quant Time: Dec 29 03:48:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D12222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 23 05:54:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	130095	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	231127	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	199825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	86864	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	66988	52.679	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.360%
35) Dibromofluoromethane	7.804	113	70896	52.128	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.260%
50) Toluene-d8	10.269	98	277398	50.479	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.960%
62) 4-Bromofluorobenzene	12.575	95	92628	52.031	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122822\
Data File : VD075060.D
Acq On : 28 Dec 2022 12:34
Operator : KP/SY
Sample : VD1228SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1228SBL01

Quant Time: Dec 29 03:48:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122222S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 23 05:54:19 2022
Response via : Initial Calibration

