

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122822\
 Data File : VD075064.D
 Acq On : 28 Dec 2022 14:27
 Operator : KP/SY
 Sample : N6220-04
 Misc : 5.72G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-CR

Quant Time: Dec 29 03:50:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D12222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 23 05:54:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	108682	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	266221	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	248211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	109851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	70971	66.807	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 133.620%	
35) Dibromofluoromethane	7.804	113	81785	52.207	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.420%	
50) Toluene-d8	10.269	98	322879	51.010	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.020%	
62) 4-Bromofluorobenzene	12.575	95	112316	54.773	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 109.540%	
Target Compounds						
16) Acetone	4.022	43	11422	41.398	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122822\
Data File : VD075064.D
Acq On : 28 Dec 2022 14:27
Operator : KP/SY
Sample : N6220-04
Misc : 5.72G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-CR

Quant Time: Dec 29 03:50:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D12222S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 23 05:54:19 2022
Response via : Initial Calibration

