

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122822\
 Data File : VD075070.D
 Acq On : 28 Dec 2022 17:15
 Operator : KP/SY
 Sample : N6218-05
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CHRT-26598

Quant Time: Dec 29 03:52:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D12222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 23 05:54:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	2662	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	6401	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	6189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	2462	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	2521	96.886	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	193.780%#
35) Dibromofluoromethane	7.804	113	2133	56.629	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.260%
50) Toluene-d8	10.269	98	8217	53.991	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.980%
62) 4-Bromofluorobenzene	12.581	95	2745	55.675	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122822\
Data File : VD075070.D
Acq On : 28 Dec 2022 17:15
Operator : KP/SY
Sample : N6218-05
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CHRT-26598

Quant Time: Dec 29 03:52:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122222S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 23 05:54:19 2022
Response via : Initial Calibration

