

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122822\
 Data File : VD075075.D
 Acq On : 28 Dec 2022 19:34
 Operator : KP/SY
 Sample : N6231-04
 Misc : 5.32G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-2S

Quant Time: Dec 29 03:54:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D12222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 23 05:54:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	109814	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	252471	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	225478	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	92789	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	67392	62.784	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 125.560%	
35) Dibromofluoromethane	7.810	113	76003	51.158	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.320%	
50) Toluene-d8	10.269	98	302397	50.376	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.760%	
62) 4-Bromofluorobenzene	12.575	95	98421	50.611	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 101.220%	
Target Compounds						
16) Acetone	4.022	43	4303	15.435	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122822\
Data File : VD075075.D
Acq On : 28 Dec 2022 19:34
Operator : KP/SY
Sample : N6231-04
Misc : 5.32G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-2S

Quant Time: Dec 29 03:54:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122222S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 23 05:54:19 2022
Response via : Initial Calibration

